

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0080519\
 Data File : P0058857.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 05 Aug 2019 12:21
 Operator : SM/AJ
 Sample : K4150-07DL 5X
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 S17-02-BDL

Manual Integrations
 APPROVED

Ankita
 8/6/2019 2:28:52 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 05 14:16:12 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072819.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jul 27 06:49:05 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.157	3.516	222408	182391	5.270	4.565
2)	SA Decachlor...	9.629	8.399	183284	155509	3.384m	3.940
Target Compounds							
26)	L6 AR-1254-1	6.148	5.356	959995	842583	406.555m	325.254
27)	L6 AR-1254-2	6.362	5.502	1771319	920351	459.611m	402.000
28)	L6 AR-1254-3	6.727	5.899	1902440	1683741	469.781m	457.618
29)	L6 AR-1254-4	7.009	6.125	1526364	932911	423.914m	401.215
30)	L6 AR-1254-5	7.423	6.534	2088639	4377906	576.214m	1269.122m#
31)	L7 AR-1260-1	6.885	6.025	994383	919113	345.374	386.174
32)	L7 AR-1260-2	7.140	6.213	1559516	902761	374.879m	298.009
33)	L7 AR-1260-3	7.492	6.361	670978	808545	180.178	299.479 #
34)	L7 AR-1260-4	7.715	6.825	956355	387771	280.639	171.616 #
35)	L7 AR-1260-5	8.023	7.068	1403159	1187392	190.782	220.486m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO080519\
Data File : PO058857.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Aug 2019 12:21
Operator : SM/AJ
Sample : K4150-07DL 5X
Misc :
ALS Vial : 14 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
S17-02-BDL


Manual Integrations
APPROVED

Ankita
8/6/2019 2:28:52 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 05 14:16:12 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072819.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jul 27 06:49:05 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.157 min
Delta R.T.: -0.022 min
Response: 222408
Conc: 5.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
S17-02-BDL


Manual Integrations
APPROVED

Ankita
8/6/2019 2:28:52 PM




#1 Tetrachloro-m-xylene

R.T.: 3.516 min
Delta R.T.: -0.009 min
Response: 182391
Conc: 4.57 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.629 min
Delta R.T.: -0.044 min
Response: 183284
Conc: 3.38 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.399 min
Delta R.T.: -0.015 min
Response: 155509
Conc: 3.94 ng/ml




#26 AR-1254-1

R.T.: 6.148 min
Delta R.T.: -0.026 min
Response: 959995
Conc: 406.56 ng/ml m

Instrument :
ECD_O
ClientSampleId :
S17-02-BDL


Manual Integrations
APPROVED

Ankita
8/6/2019 2:28:52 PM




#26 AR-1254-1

R.T.: 5.356 min
Delta R.T.: -0.012 min
Response: 842583
Conc: 325.25 ng/ml




#27 AR-1254-2

R.T.: 6.362 min
Delta R.T.: -0.027 min
Response: 1771319
Conc: 459.61 ng/ml m



#27 AR-1254-2

R.T.: 5.502 min
Delta R.T.: -0.012 min
Response: 920351
Conc: 402.00 ng/ml




#28 AR-1254-3

R.T.: 6.727 min
Delta R.T.: -0.027 min
Response: 1902440
Conc: 469.78 ng/ml m

Instrument :
ECD_O
ClientSampleId :
S17-02-BDL


Manual Integrations
APPROVED

Ankita
8/6/2019 2:28:52 PM




#28 AR-1254-3

R.T.: 5.899 min
Delta R.T.: -0.012 min
Response: 1683741
Conc: 457.62 ng/ml




#29 AR-1254-4

R.T.: 7.009 min
Delta R.T.: -0.027 min
Response: 1526364
Conc: 423.91 ng/ml m



#29 AR-1254-4

R.T.: 6.125 min
Delta R.T.: -0.013 min
Response: 932911
Conc: 401.22 ng/ml




#30 AR-1254-5

R.T.: 7.423 min
Delta R.T.: -0.028 min
Response: 2088639
Conc: 576.21 ng/ml m

Instrument :
ECD_O
ClientSampleId :
S17-02-BDL


Manual Integrations
APPROVED

Ankita
8/6/2019 2:28:52 PM




#30 AR-1254-5

R.T.: 6.534 min
Delta R.T.: -0.014 min
Response: 4377906
Conc: 1269.12 ng/ml m




#31 AR-1260-1

R.T.: 6.885 min
Delta R.T.: -0.026 min
Response: 994383
Conc: 345.37 ng/ml



#31 AR-1260-1

R.T.: 6.025 min
Delta R.T.: -0.012 min
Response: 919113
Conc: 386.17 ng/ml




#32 AR-1260-2

R.T.: 7.140 min
Delta R.T.: -0.027 min
Response: 1559516
Conc: 374.88 ng/ml m

Instrument :
ECD_O
ClientSampleId :
S17-02-BDL


Manual Integrations
APPROVED

Ankita
8/6/2019 2:28:52 PM




#32 AR-1260-2

R.T.: 6.213 min
Delta R.T.: -0.012 min
Response: 902761
Conc: 298.01 ng/ml




#33 AR-1260-3

R.T.: 7.492 min
Delta R.T.: -0.028 min
Response: 670978
Conc: 180.18 ng/ml



#33 AR-1260-3

R.T.: 6.361 min
Delta R.T.: -0.013 min
Response: 808545
Conc: 299.48 ng/ml




#34 AR-1260-4

R.T.: 7.715 min
Delta R.T.: -0.031 min
Response: 956355
Conc: 280.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
S17-02-BDL


Manual Integrations
APPROVED

Ankita
8/6/2019 2:28:52 PM




#34 AR-1260-4

R.T.: 6.825 min
Delta R.T.: -0.014 min
Response: 387771
Conc: 171.62 ng/ml



#35 AR-1260-5

R.T.: 8.023 min
Delta R.T.: -0.028 min
Response: 1403159
Conc: 190.78 ng/ml



#35 AR-1260-5

R.T.: 7.068 min
Delta R.T.: -0.014 min
Response: 1187392
Conc: 220.49 ng/ml m