

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0080521\
 Data File : P0080228.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 05 Aug 2021 22:19
 Operator : DD\AJ
 Sample : PB138220BS
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB138220BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 06 00:26:10 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 28 18:45:36 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.996	4.113	1229137	465699	20.398	20.335
2)	SA Decachlor...	11.100	9.518	925291	442515	19.725	21.294
Target Compounds							
3)	L1 AR-1016-1	6.326	5.389	135633	65061	61.701	69.219
4)	L1 AR-1016-2	6.350	5.410	187020	88945	61.297	69.914
5)	L1 AR-1016-3	6.416	5.605	118308	46522	62.869	65.987
6)	L1 AR-1016-4	6.524	5.656	92708	42101	60.126	69.673
7)	L1 AR-1016-5	6.840	5.889	91447	48061	60.444	65.260
31)	L7 AR-1260-1	8.021	7.000	162998	99478	66.436	76.854
32)	L7 AR-1260-2	8.287	7.199	213401	120786	69.440	78.957
33)	L7 AR-1260-3	8.655	7.357	122986	117652	55.466	80.580 #
34)	L7 AR-1260-4	8.887	7.846	146024	78969	56.884	66.643
35)	L7 AR-1260-5	9.221	8.096	281023	163894	54.963	62.042



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO080521\
Data File : PO080228.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Aug 2021 22:19
Operator : DD\AJ
Sample : PB138220BS
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB138220BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 06 00:26:10 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072821.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 28 18:45:36 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.996 min
Delta R.T.: -0.004 min
Response: 1229137
Conc: 20.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB138220BS




#1 Tetrachloro-m-xylene

R.T.: 4.113 min
Delta R.T.: -0.006 min
Response: 465699
Conc: 20.34 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.100 min
Delta R.T.: -0.010 min
Response: 925291
Conc: 19.72 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.518 min
Delta R.T.: -0.012 min
Response: 442515
Conc: 21.29 ng/ml



#3 AR-1016-1


R.T.: 6.326 min
Delta R.T.: -0.005 min
Response: 135633
Conc: 61.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB138220BS




#3 AR-1016-1

R.T.: 5.389 min
Delta R.T.: -0.010 min
Response: 65061
Conc: 69.22 ng/ml




#4 AR-1016-2

R.T.: 6.350 min
Delta R.T.: -0.006 min
Response: 187020
Conc: 61.30 ng/ml



#4 AR-1016-2


R.T.: 5.410 min
Delta R.T.: -0.009 min
Response: 88945
Conc: 69.91 ng/ml



#5 AR-1016-3


R.T.: 6.416 min
Delta R.T.: -0.006 min
Response: 118308
Conc: 62.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB138220BS




#5 AR-1016-3

R.T.: 5.605 min
Delta R.T.: -0.009 min
Response: 46522
Conc: 65.99 ng/ml




#6 AR-1016-4

R.T.: 6.524 min
Delta R.T.: -0.005 min
Response: 92708
Conc: 60.13 ng/ml



#6 AR-1016-4


R.T.: 5.656 min
Delta R.T.: -0.008 min
Response: 42101
Conc: 69.67 ng/ml



#7 AR-1016-5


R.T.: 6.840 min
Delta R.T.: -0.005 min
Response: 91447
Conc: 60.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB138220BS




#7 AR-1016-5

R.T.: 5.889 min
Delta R.T.: -0.010 min
Response: 48061
Conc: 65.26 ng/ml




#31 AR-1260-1

R.T.: 8.021 min
Delta R.T.: -0.006 min
Response: 162998
Conc: 66.44 ng/ml



#31 AR-1260-1


R.T.: 7.000 min
Delta R.T.: -0.010 min
Response: 99478
Conc: 76.85 ng/ml



#32 AR-1260-2


R.T.: 8.287 min
Delta R.T.: -0.006 min
Response: 213401
Conc: 69.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB138220BS




#32 AR-1260-2

R.T.: 7.199 min
Delta R.T.: -0.010 min
Response: 120786
Conc: 78.96 ng/ml




#33 AR-1260-3

R.T.: 8.655 min
Delta R.T.: -0.006 min
Response: 122986
Conc: 55.47 ng/ml



#33 AR-1260-3


R.T.: 7.357 min
Delta R.T.: -0.010 min
Response: 117652
Conc: 80.58 ng/ml



#34 AR-1260-4


R.T.: 8.887 min
Delta R.T.: -0.006 min
Response: 146024
Conc: 56.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB138220BS




#34 AR-1260-4

R.T.: 7.846 min
Delta R.T.: -0.011 min
Response: 78969
Conc: 66.64 ng/ml



#35 AR-1260-5

R.T.: 9.221 min
Delta R.T.: -0.006 min
Response: 281023
Conc: 54.96 ng/ml



#35 AR-1260-5

R.T.: 8.096 min
Delta R.T.: -0.010 min
Response: 163894
Conc: 62.04 ng/ml