

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0080522\
 Data File : P0088636.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Aug 2022 02:22
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 06 07:30:00 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 05 23:24:09 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.437	3.594	143.5E6	55453040	53.215	51.236
2)	SA Decachlor...	10.274	8.586	94987284	55542146	49.224	45.830
Target Compounds							
21)	L5 AR-1248-1	5.612	4.673	27859448	12419531	539.998	528.923
22)	L5 AR-1248-2	5.888	4.909	40982601	18706803	533.850	521.239
23)	L5 AR-1248-3	6.094	4.950	44464401	20223947	538.315	531.147
24)	L5 AR-1248-4	6.499	5.121	48098211	24268091	535.569	528.457
25)	L5 AR-1248-5	6.538	5.513	48296188	23212567	544.848	531.657



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO080522\
Data File : PO088636.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Aug 2022 02:22
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 06 07:30:00 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO080422.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 05 23:24:09 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.437 min
Delta R.T.: -0.002 min
Response: 143539807
Conc: 53.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.594 min
Delta R.T.: -0.003 min
Response: 55453040
Conc: 51.24 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.274 min
Delta R.T.: -0.003 min
Response: 94987284
Conc: 49.22 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.586 min
Delta R.T.: -0.005 min
Response: 55542146
Conc: 45.83 ng/ml



#21 AR-1248-1


R.T.: 5.612 min
Delta R.T.: -0.001 min
Response: 27859448
Conc: 540.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.673 min
Delta R.T.: -0.002 min
Response: 12419531
Conc: 528.92 ng/ml




#22 AR-1248-2

R.T.: 5.888 min
Delta R.T.: -0.002 min
Response: 40982601
Conc: 533.85 ng/ml



#22 AR-1248-2


R.T.: 4.909 min
Delta R.T.: -0.002 min
Response: 18706803
Conc: 521.24 ng/ml



#23 AR-1248-3


R.T.: 6.094 min
Delta R.T.: -0.001 min
Response: 44464401
Conc: 538.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.950 min
Delta R.T.: -0.002 min
Response: 20223947
Conc: 531.15 ng/ml




#24 AR-1248-4

R.T.: 6.499 min
Delta R.T.: -0.003 min
Response: 48098211
Conc: 535.57 ng/ml



#24 AR-1248-4


R.T.: 5.121 min
Delta R.T.: -0.003 min
Response: 24268091
Conc: 528.46 ng/ml



#25 AR-1248-5

R.T.: 6.538 min
Delta R.T.: -0.002 min
Response: 48296188
Conc: 544.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.513 min
Delta R.T.: -0.002 min
Response: 23212567
Conc: 531.66 ng/ml