

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0080522\
 Data File : P0088663.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Aug 2022 10:48
 Operator : YP/AJ
 Sample : N4045-05
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 40058

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 06 14:29:29 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 05 23:24:09 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.435	3.593	66552975	23790976	24.673	21.982
2)	SA Decachlor...	10.274	8.587	47243227	27901371	24.482	23.023

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO080522\
Data File : PO088663.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Aug 2022 10:48
Operator : YP/AJ
Sample : N4045-05
Misc :
ALS Vial : 45 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
40058

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 06 14:29:29 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO080422.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 05 23:24:09 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.435 min
Delta R.T.: -0.004 min
Response: 66552975
Conc: 24.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
40058




#1 Tetrachloro-m-xylene

R.T.: 3.593 min
Delta R.T.: -0.005 min
Response: 23790976
Conc: 21.98 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.274 min
Delta R.T.: -0.003 min
Response: 47243227
Conc: 24.48 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.587 min
Delta R.T.: -0.005 min
Response: 27901371
Conc: 23.02 ng/ml