

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0080522\
 Data File : P0088669.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Aug 2022 13:13
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/08/2022
 Supervised By :Ankita Jodhani 08/08/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 06 14:35:00 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 05 23:24:09 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.436	3.593	150.7E6	54944174	55.886	50.766
2) SA Decachlor...	10.275	8.588	102.5E6	60330924	53.118	49.782
Target Compounds						
26) L6 AR-1254-1	6.474	5.472	54775729	36225828	599.644m	525.969
27) L6 AR-1254-2	6.693	5.619	81855625	31662871	598.564m	534.664
28) L6 AR-1254-3	7.062	6.023	84309967	48264046	599.106m	519.033
29) L6 AR-1254-4	7.349	6.251	55562158	29625334	549.875m	522.948
30) L6 AR-1254-5	7.770	6.669	58942217	43295892	553.504	530.411

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO080522\
Data File : PO088669.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Aug 2022 13:13
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/08/2022
Supervised By :Ankita Jodhani 08/08/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 06 14:35:00 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO080422.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 05 23:24:09 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.436 min
Delta R.T.: -0.003 min
Response: 150745689
Conc: 55.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/08/2022
Supervised By :Ankita Jodhani 08/08/2022




#1 Tetrachloro-m-xylene

R.T.: 3.593 min
Delta R.T.: -0.004 min
Response: 54944174
Conc: 50.77 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.275 min
Delta R.T.: -0.002 min
Response: 102501865
Conc: 53.12 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.588 min
Delta R.T.: -0.004 min
Response: 60330924
Conc: 49.78 ng/ml




#26 AR-1254-1

R.T.: 6.474 min
Delta R.T.: 0.000 min
Response: 54775729
Conc: 599.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/08/2022
Supervised By :Ankita Jodhani 08/08/2022




#26 AR-1254-1

R.T.: 5.472 min
Delta R.T.: -0.003 min
Response: 36225828
Conc: 525.97 ng/ml




#27 AR-1254-2

R.T.: 6.693 min
Delta R.T.: 0.000 min
Response: 81855625
Conc: 598.56 ng/ml m



#27 AR-1254-2

R.T.: 5.619 min
Delta R.T.: -0.003 min
Response: 31662871
Conc: 534.66 ng/ml




#28 AR-1254-3

R.T.: 7.062 min
Delta R.T.: 0.000 min
Response: 84309967
Conc: 599.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/08/2022
Supervised By :Ankita Jodhani 08/08/2022




#28 AR-1254-3

R.T.: 6.023 min
Delta R.T.: -0.002 min
Response: 48264046
Conc: 519.03 ng/ml




#29 AR-1254-4

R.T.: 7.349 min
Delta R.T.: 0.000 min
Response: 55562158
Conc: 549.88 ng/ml m



#29 AR-1254-4

R.T.: 6.251 min
Delta R.T.: -0.002 min
Response: 29625334
Conc: 522.95 ng/ml




#30 AR-1254-5

R.T.: 7.770 min
Delta R.T.: 0.000 min
Response: 58942217
Conc: 553.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/08/2022
Supervised By :Ankita Jodhani 08/08/2022



#30 AR-1254-5

R.T.: 6.669 min
Delta R.T.: -0.002 min
Response: 43295892
Conc: 530.41 ng/ml