

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0080620\
 Data File : P0070354.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Aug 2020 2:12
 Operator : DD\AJ
 Sample : PB130781BS
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB130781BSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 07 05:36:36 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080620.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Aug 06 13:04:06 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.375	3.561	1388753	803736	21.002	18.242
2)	SA Decachlor...	10.004	8.753	1881807	1189070	21.977	21.422
Target Compounds							
3)	L1 AR-1016-1	5.680	4.790	604241	375687	209.268	186.577
4)	L1 AR-1016-2	5.701	4.808	863695	513126	208.854	180.575
5)	L1 AR-1016-3	5.765	4.993	516151	281818	208.894	184.578
6)	L1 AR-1016-4	5.871	5.050	439161	238066	209.618	183.000
7)	L1 AR-1016-5	6.179	5.270	429408	298584	205.682	184.910
8)	L2 AR-1221-1	4.625	3.812	45668	30250	54.468	49.863
9)	L2 AR-1221-2	4.726	3.908	71432	43244	113.544	97.736
10)	L2 AR-1221-3	4.812	3.994	254290	161789	127.525	117.891
11)	L3 AR-1232-1	4.812	3.994	254290	161789	168.993	149.146
12)	L3 AR-1232-2	5.387	4.808	387756	513126	471.337	413.307
13)	L3 AR-1232-3	5.701	4.993	863695	281818	484.364	422.402
14)	L3 AR-1232-4	5.871	5.091	439161	257061	481.342	467.170
15)	L3 AR-1232-5	5.969	5.270	346784	298584	567.094	454.082
16)	L4 AR-1242-1	5.680	4.790	604241	375687	256.519	229.505
17)	L4 AR-1242-2	5.701	4.808	863695	513126	256.000	222.644
18)	L4 AR-1242-3	5.765	4.993	516151	281818	252.179	226.471
19)	L4 AR-1242-4	5.871	5.091	439161	257061	253.416	226.624
20)	L4 AR-1242-5	6.641	5.645	107792	249553	52.331	149.609 #
21)	L5 AR-1248-1	5.680	4.790	604241	375687	339.918	315.620
22)	L5 AR-1248-2	5.969	5.050	346784	238066	152.907	139.677
23)	L5 AR-1248-3	6.179	5.091	429408	257061	156.761	163.860
24)	L5 AR-1248-4	6.606	5.270	312983	298584	91.988	146.861 #
25)	L5 AR-1248-5	6.641	5.687	107792	31618	33.888	15.340 #
26)	L6 AR-1254-1	6.571	5.645	496992	249553	139.665	71.260 #
27)	L6 AR-1254-2	6.799	5.806	400296	286352	71.893	91.947 #
28)	L6 AR-1254-3	7.175	6.217	213516	132689	37.076	26.916 #
29)	L6 AR-1254-4	7.466	6.457	157621	65321	30.216	18.149 #
30)	L6 AR-1254-5	7.890	6.878	1523429	1016163	288.368	220.929
31)	L7 AR-1260-1	7.337	6.353	967979	677986	232.740	209.964
32)	L7 AR-1260-2	7.603	6.553	1331657	926802	221.778	215.391
33)	L7 AR-1260-3	7.961	6.701	959963	787260	186.517	215.018
34)	L7 AR-1260-4	8.188	7.177	955963	603461	196.593	193.576
35)	L7 AR-1260-5	8.502	7.429	2099301	1559221	188.344	192.980
36)	L8 AR-1262-1	7.961	6.878	959963	1016163	125.333	401.973 #
37)	L8 AR-1262-2	8.502	7.429	2099301	1559221	166.635	170.214
38)	L8 AR-1262-3	8.770	7.710	391039	312724	68.300	85.458 #
39)	L8 AR-1262-4	8.835f	7.770	690732	1034312	218.571	164.023
40)	L8 AR-1262-5	9.407	8.269	449612	305255	108.807	106.360
41)	L9 AR-1268-1	8.770	7.710	391039	312724	24.670	27.843
42)	L9 AR-1268-2	8.835f	7.770	690732	1034312	50.777	107.222 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0080620\
 Data File : P0070354.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Aug 2020 2:12
 Operator : DD\AJ
 Sample : PB130781BS
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB130781BSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 07 05:36:36 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080620.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Aug 06 13:04:06 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
43) L9	AR-1268-3	9.034	7.975	40550	28702	3.484	3.569
44) L9	AR-1268-4	9.407	8.269	449612	305255	93.189	92.847
45) L9	AR-1268-5	9.739	8.537	149254	104503	4.751	4.773



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO080620\
Data File : PO070354.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 07 Aug 2020 2:12
Operator : DD\AJ
Sample : PB130781BS
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
PB130781BSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 07 05:36:36 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO080620.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Aug 06 13:04:06 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.375 min
Delta R.T.: 0.001 min
Response: 1388753
Conc: 21.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB130781BSD




#1 Tetrachloro-m-xylene

R.T.: 3.561 min
Delta R.T.: 0.002 min
Response: 803736
Conc: 18.24 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.004 min
Delta R.T.: -0.001 min
Response: 1881807
Conc: 21.98 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.753 min
Delta R.T.: 0.000 min
Response: 1189070
Conc: 21.42 ng/ml



#3 AR-1016-1


R.T.: 5.680 min
Delta R.T.: 0.000 min
Response: 604241
Conc: 209.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB130781BSD




#3 AR-1016-1

R.T.: 4.790 min
Delta R.T.: 0.002 min
Response: 375687
Conc: 186.58 ng/ml




#4 AR-1016-2

R.T.: 5.701 min
Delta R.T.: 0.000 min
Response: 863695
Conc: 208.85 ng/ml



#4 AR-1016-2


R.T.: 4.808 min
Delta R.T.: 0.000 min
Response: 513126
Conc: 180.57 ng/ml



#5 AR-1016-3


R.T.: 5.765 min
Delta R.T.: 0.000 min
Response: 516151
Conc: 208.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB130781BSD




#5 AR-1016-3

R.T.: 4.993 min
Delta R.T.: 0.002 min
Response: 281818
Conc: 184.58 ng/ml




#6 AR-1016-4

R.T.: 5.871 min
Delta R.T.: 0.000 min
Response: 439161
Conc: 209.62 ng/ml



#6 AR-1016-4


R.T.: 5.050 min
Delta R.T.: 0.000 min
Response: 238066
Conc: 183.00 ng/ml



#7 AR-1016-5


R.T.: 6.179 min
Delta R.T.: 0.000 min
Response: 429408
Conc: 205.68 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB130781BSD




#7 AR-1016-5

R.T.: 5.270 min
Delta R.T.: 0.000 min
Response: 298584
Conc: 184.91 ng/ml




#8 AR-1221-1

R.T.: 4.625 min
Delta R.T.: -0.006 min
Response: 45668
Conc: 54.47 ng/ml



#8 AR-1221-1


R.T.: 3.812 min
Delta R.T.: 0.000 min
Response: 30250
Conc: 49.86 ng/ml



#9 AR-1221-2


R.T.: 4.726 min
Delta R.T.: 0.000 min
Response: 71432
Conc: 113.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB130781BSD




#9 AR-1221-2

R.T.: 3.908 min
Delta R.T.: 0.000 min
Response: 43244
Conc: 97.74 ng/ml




#10 AR-1221-3

R.T.: 4.812 min
Delta R.T.: 0.000 min
Response: 254290
Conc: 127.52 ng/ml



#10 AR-1221-3


R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 161789
Conc: 117.89 ng/ml



#11 AR-1232-1


R.T.: 4.812 min
Delta R.T.: 0.001 min
Response: 254290
Conc: 168.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB130781BSD




#11 AR-1232-1

R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 161789
Conc: 149.15 ng/ml




#12 AR-1232-2

R.T.: 5.387 min
Delta R.T.: 0.000 min
Response: 387756
Conc: 471.34 ng/ml



#12 AR-1232-2


R.T.: 4.808 min
Delta R.T.: 0.002 min
Response: 513126
Conc: 413.31 ng/ml



#13 AR-1232-3


R.T.: 5.701 min
Delta R.T.: 0.000 min
Response: 863695
Conc: 484.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB130781BSD




#13 AR-1232-3

R.T.: 4.993 min
Delta R.T.: 0.001 min
Response: 281818
Conc: 422.40 ng/ml




#14 AR-1232-4

R.T.: 5.871 min
Delta R.T.: 0.000 min
Response: 439161
Conc: 481.34 ng/ml



#14 AR-1232-4


R.T.: 5.091 min
Delta R.T.: 0.001 min
Response: 257061
Conc: 467.17 ng/ml



#15 AR-1232-5


R.T.: 5.969 min
Delta R.T.: 0.000 min
Response: 346784
Conc: 567.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB130781BSD




#15 AR-1232-5

R.T.: 5.270 min
Delta R.T.: 0.000 min
Response: 298584
Conc: 454.08 ng/ml




#16 AR-1242-1

R.T.: 5.680 min
Delta R.T.: 0.001 min
Response: 604241
Conc: 256.52 ng/ml



#16 AR-1242-1


R.T.: 4.790 min
Delta R.T.: 0.002 min
Response: 375687
Conc: 229.51 ng/ml



#17 AR-1242-2


R.T.: 5.701 min
Delta R.T.: 0.001 min
Response: 863695
Conc: 256.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB130781BSD




#17 AR-1242-2

R.T.: 4.808 min
Delta R.T.: 0.002 min
Response: 513126
Conc: 222.64 ng/ml




#18 AR-1242-3

R.T.: 5.765 min
Delta R.T.: 0.000 min
Response: 516151
Conc: 252.18 ng/ml



#18 AR-1242-3


R.T.: 4.993 min
Delta R.T.: 0.001 min
Response: 281818
Conc: 226.47 ng/ml



#19 AR-1242-4


R.T.: 5.871 min
Delta R.T.: 0.001 min
Response: 439161
Conc: 253.42 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB130781BSD




#19 AR-1242-4

R.T.: 5.091 min
Delta R.T.: 0.000 min
Response: 257061
Conc: 226.62 ng/ml




#20 AR-1242-5

R.T.: 6.641 min
Delta R.T.: 0.000 min
Response: 107792
Conc: 52.33 ng/ml



#20 AR-1242-5


R.T.: 5.645 min
Delta R.T.: 0.000 min
Response: 249553
Conc: 149.61 ng/ml



#21 AR-1248-1


R.T.: 5.680 min
Delta R.T.: 0.000 min
Response: 604241
Conc: 339.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB130781BSD




#21 AR-1248-1

R.T.: 4.790 min
Delta R.T.: 0.002 min
Response: 375687
Conc: 315.62 ng/ml




#22 AR-1248-2

R.T.: 5.969 min
Delta R.T.: 0.000 min
Response: 346784
Conc: 152.91 ng/ml



#22 AR-1248-2


R.T.: 5.050 min
Delta R.T.: 0.001 min
Response: 238066
Conc: 139.68 ng/ml



#23 AR-1248-3


R.T.: 6.179 min
Delta R.T.: 0.000 min
Response: 429408
Conc: 156.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB130781BSD




#23 AR-1248-3

R.T.: 5.091 min
Delta R.T.: 0.001 min
Response: 257061
Conc: 163.86 ng/ml




#24 AR-1248-4

R.T.: 6.606 min
Delta R.T.: 0.002 min
Response: 312983
Conc: 91.99 ng/ml



#24 AR-1248-4


R.T.: 5.270 min
Delta R.T.: 0.001 min
Response: 298584
Conc: 146.86 ng/ml



#25 AR-1248-5


R.T.: 6.641 min
Delta R.T.: 0.000 min
Response: 107792
Conc: 33.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB130781BSD




#25 AR-1248-5

R.T.: 5.687 min
Delta R.T.: 0.000 min
Response: 31618
Conc: 15.34 ng/ml




#26 AR-1254-1

R.T.: 6.571 min
Delta R.T.: 0.002 min
Response: 496992
Conc: 139.66 ng/ml



#26 AR-1254-1


R.T.: 5.645 min
Delta R.T.: 0.002 min
Response: 249553
Conc: 71.26 ng/ml



#27 AR-1254-2


R.T.: 6.799 min
Delta R.T.: 0.000 min
Response: 400296
Conc: 71.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB130781BSD




#27 AR-1254-2

R.T.: 5.806 min
Delta R.T.: 0.002 min
Response: 286352
Conc: 91.95 ng/ml




#28 AR-1254-3

R.T.: 7.175 min
Delta R.T.: 0.001 min
Response: 213516
Conc: 37.08 ng/ml



#28 AR-1254-3


R.T.: 6.217 min
Delta R.T.: 0.003 min
Response: 132689
Conc: 26.92 ng/ml



#29 AR-1254-4


R.T.: 7.466 min
Delta R.T.: -0.004 min
Response: 157621
Conc: 30.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB130781BSD




#29 AR-1254-4

R.T.: 6.457 min
Delta R.T.: 0.001 min
Response: 65321
Conc: 18.15 ng/ml




#30 AR-1254-5

R.T.: 7.890 min
Delta R.T.: 0.000 min
Response: 1523429
Conc: 288.37 ng/ml



#30 AR-1254-5


R.T.: 6.878 min
Delta R.T.: 0.000 min
Response: 1016163
Conc: 220.93 ng/ml



#31 AR-1260-1


R.T.: 7.337 min
Delta R.T.: -0.001 min
Response: 967979
Conc: 232.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB130781BSD




#31 AR-1260-1

R.T.: 6.353 min
Delta R.T.: 0.000 min
Response: 677986
Conc: 209.96 ng/ml




#32 AR-1260-2

R.T.: 7.603 min
Delta R.T.: 0.000 min
Response: 1331657
Conc: 221.78 ng/ml



#32 AR-1260-2


R.T.: 6.553 min
Delta R.T.: 0.000 min
Response: 926802
Conc: 215.39 ng/ml



#33 AR-1260-3


R.T.: 7.961 min
Delta R.T.: 0.000 min
Response: 959963
Conc: 186.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB130781BSD




#33 AR-1260-3

R.T.: 6.701 min
Delta R.T.: 0.000 min
Response: 787260
Conc: 215.02 ng/ml




#34 AR-1260-4

R.T.: 8.188 min
Delta R.T.: -0.001 min
Response: 955963
Conc: 196.59 ng/ml



#34 AR-1260-4


R.T.: 7.177 min
Delta R.T.: 0.000 min
Response: 603461
Conc: 193.58 ng/ml



#35 AR-1260-5


R.T.: 8.502 min
Delta R.T.: -0.001 min
Response: 2099301
Conc: 188.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB130781BSD




#35 AR-1260-5

R.T.: 7.429 min
Delta R.T.: 0.000 min
Response: 1559221
Conc: 192.98 ng/ml




#36 AR-1262-1

R.T.: 7.961 min
Delta R.T.: 0.000 min
Response: 959963
Conc: 125.33 ng/ml



#36 AR-1262-1


R.T.: 6.878 min
Delta R.T.: 0.002 min
Response: 1016163
Conc: 401.97 ng/ml



#37 AR-1262-2


R.T.: 8.502 min
Delta R.T.: 0.000 min
Response: 2099301
Conc: 166.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB130781BSD




#37 AR-1262-2

R.T.: 7.429 min
Delta R.T.: 0.001 min
Response: 1559221
Conc: 170.21 ng/ml




#38 AR-1262-3

R.T.: 8.770 min
Delta R.T.: 0.000 min
Response: 391039
Conc: 68.30 ng/ml



#38 AR-1262-3


R.T.: 7.710 min
Delta R.T.: 0.000 min
Response: 312724
Conc: 85.46 ng/ml



#39 AR-1262-4


R.T.: 8.835 min
Delta R.T.: -0.016 min
Response: 690732
Conc: 218.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB130781BSD




#39 AR-1262-4

R.T.: 7.770 min
Delta R.T.: -0.002 min
Response: 1034312
Conc: 164.02 ng/ml




#40 AR-1262-5

R.T.: 9.407 min
Delta R.T.: -0.001 min
Response: 449612
Conc: 108.81 ng/ml



#40 AR-1262-5


R.T.: 8.269 min
Delta R.T.: 0.001 min
Response: 305255
Conc: 106.36 ng/ml



#41 AR-1268-1


R.T.: 8.770 min
Delta R.T.: 0.000 min
Response: 391039
Conc: 24.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB130781BSD




#41 AR-1268-1

R.T.: 7.710 min
Delta R.T.: 0.000 min
Response: 312724
Conc: 27.84 ng/ml




#42 AR-1268-2

R.T.: 8.835 min
Delta R.T.: -0.017 min
Response: 690732
Conc: 50.78 ng/ml



#42 AR-1268-2


R.T.: 7.770 min
Delta R.T.: -0.004 min
Response: 1034312
Conc: 107.22 ng/ml



#43 AR-1268-3


R.T.: 9.034 min
Delta R.T.: 0.000 min
Response: 40550
Conc: 3.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB130781BSD




#43 AR-1268-3

R.T.: 7.975 min
Delta R.T.: 0.000 min
Response: 28702
Conc: 3.57 ng/ml




#44 AR-1268-4

R.T.: 9.407 min
Delta R.T.: 0.000 min
Response: 449612
Conc: 93.19 ng/ml



#44 AR-1268-4


R.T.: 8.269 min
Delta R.T.: 0.001 min
Response: 305255
Conc: 92.85 ng/ml



#45 AR-1268-5

R.T.: 9.739 min
Delta R.T.: 0.000 min
Response: 149254
Conc: 4.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB130781BSD



#45 AR-1268-5

R.T.: 8.537 min
Delta R.T.: 0.000 min
Response: 104503
Conc: 4.77 ng/ml