

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_0\Data\P0080620\  
 Data File : P0070360.D  
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
 Acq On : 07 Aug 2020 5:20  
 Operator : DD\AJ  
 Sample : L3573-01MS  
 Misc :  
 ALS Vial : 43 Sample Multiplier: 1

**Instrument :**  
 ECD\_0  
**ClientSampleId :**  
 FK-GF-03-016-AMS

**Manual Integrations**  
**APPROVED**  
 Ankita  
 8/7/2020 1:14:28 PM

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Aug 07 08:19:31 2020  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_0\methods\P0080620.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Thu Aug 06 13:04:06 2020  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
-----						
System Monitoring Compounds						
1) SA Tetrachlo...	4.376	3.562	1920154	1277067	29.038m	28.985
2) SA Decachlor...	10.005	8.754	1548548	919735	18.085	16.569
Target Compounds						
3) L1 AR-1016-1	5.682	4.790	759994	487011	263.210	241.863
4) L1 AR-1016-2	5.703	4.808	1087833	838884	263.053	295.213
5) L1 AR-1016-3	5.765	4.993	658417	442882	266.472m	290.067
6) L1 AR-1016-4	5.871	5.050	561555	327201	268.038m	251.518
7) L1 AR-1016-5	6.181	5.271	376774	437162	180.471m	270.730 #
31) L7 AR-1260-1	7.338	6.353	930198	723090	223.655	223.932
32) L7 AR-1260-2	7.609	6.553	2873395	1057526	478.543m	245.772 #
33) L7 AR-1260-3	7.963	6.700	874833	787485	169.976	215.079m#
34) L7 AR-1260-4	8.189	7.177	890894	527830	183.212	169.315
35) L7 AR-1260-5	8.504	7.429	1867534	1214777	167.551	150.349

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_O\Data\PO080620\  
 Data File : PO070360.D  
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
 Acq On : 07 Aug 2020 5:20  
 Operator : DD\AJ  
 Sample : L3573-01MS  
 Misc :  
 ALS Vial : 43 Sample Multiplier: 1

**Instrument :**  
 ECD\_O  
**Client Sampled :**  
 FK-GF-03-016-AMS

**Manual Integrations**  
**APPROVED**  
 Ankita  
 8/7/2020 1:14:28 PM

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Aug 07 08:19:31 2020  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_O\methods\PO080620.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Thu Aug 06 13:04:06 2020  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

