

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0080624\
 Data File : P0105241.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Aug 2024 19:28
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 07 02:03:52 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 30 07:11:20 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.430	3.539	406.5E6	152.7E6	59.048	59.909
2) SA Decachlor...	10.221	8.430	275.8E6	85683560	53.725	53.939
Target Compounds						
26) L6 AR-1254-1	6.462	5.378	146.9E6	79192316	507.599	527.774
27) L6 AR-1254-2	6.682	5.523	211.6E6	69435344	509.459	518.731
28) L6 AR-1254-3	7.052	5.920	218.6E6	107.2E6	523.536	504.610
29) L6 AR-1254-4	7.338	6.145	161.9E6	67078161	517.915	505.058
30) L6 AR-1254-5	7.761	6.556	177.4E6	94092788	484.904	506.904



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO080624\
Data File : PO105241.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Aug 2024 19:28
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 07 02:03:52 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 30 07:11:20 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.430 min
Delta R.T.: 0.000 min
Response: 406506388
Conc: 59.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.539 min
Delta R.T.: -0.001 min
Response: 152653757
Conc: 59.91 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.221 min
Delta R.T.: 0.000 min
Response: 275821792
Conc: 53.73 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.430 min
Delta R.T.: -0.007 min
Response: 85683560
Conc: 53.94 ng/ml



#26 AR-1254-1


R.T.: 6.462 min
Delta R.T.: 0.002 min
Response: 146913930
Conc: 507.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.378 min
Delta R.T.: -0.003 min
Response: 79192316
Conc: 527.77 ng/ml




#27 AR-1254-2

R.T.: 6.682 min
Delta R.T.: 0.001 min
Response: 211552783
Conc: 509.46 ng/ml



#27 AR-1254-2


R.T.: 5.523 min
Delta R.T.: -0.003 min
Response: 69435344
Conc: 518.73 ng/ml



#28 AR-1254-3


R.T.: 7.052 min
Delta R.T.: 0.002 min
Response: 218568916
Conc: 523.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.920 min
Delta R.T.: -0.004 min
Response: 107214885
Conc: 504.61 ng/ml




#29 AR-1254-4

R.T.: 7.338 min
Delta R.T.: 0.001 min
Response: 161869249
Conc: 517.91 ng/ml



#29 AR-1254-4


R.T.: 6.145 min
Delta R.T.: -0.003 min
Response: 67078161
Conc: 505.06 ng/ml



#30 AR-1254-5

R.T.: 7.761 min
Delta R.T.: 0.001 min
Response: 177415859
Conc: 484.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.556 min
Delta R.T.: -0.005 min
Response: 94092788
Conc: 506.90 ng/ml