

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0080724\
 Data File : P0105257.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Aug 2024 15:04
 Operator : YP/AJ
 Sample : P3487-02
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 0726-B

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/08/2024
 Supervised By :Ankita Jodhani 08/08/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 07 16:51:11 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 30 07:11:20 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.432	3.545	87422111	33548058	12.699m	13.166
2) SA Decachlor...	10.219	8.429	119.8E6	28250389	23.329	17.784

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO080724\
Data File : PO105257.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Aug 2024 15:04
Operator : YP/AJ
Sample : P3487-02
Misc :
ALS Vial : 15 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
0726-B


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/08/2024
Supervised By :Ankita Jodhani 08/08/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 07 16:51:11 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 30 07:11:20 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.432 min
Delta R.T.: 0.003 min
Response: 87422111
Conc: 12.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
0726-B


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/08/2024
Supervised By :Ankita Jodhani 08/08/2024




#1 Tetrachloro-m-xylene

R.T.: 3.545 min
Delta R.T.: 0.005 min
Response: 33548058
Conc: 13.17 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.219 min
Delta R.T.: -0.002 min
Response: 119771785
Conc: 23.33 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.429 min
Delta R.T.: -0.008 min
Response: 28250389
Conc: 17.78 ng/ml