

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0080724\
 Data File : P0105264.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Aug 2024 17:48
 Operator : YP/AJ
 Sample : P3491-03 5X
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 COMP-1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 08 02:41:11 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 30 07:11:20 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.433	3.546	18149540	6737124	2.636	2.644
2) SA Decachlor...	10.220	8.428	26289524	7293908	5.121	4.592
Target Compounds						
31) L7 AR-1260-1	7.218	6.051	374.1E6	133.1E6	1178.179	977.184
32) L7 AR-1260-2	7.477	6.236	463.2E6	163.5E6	1273.757	971.770
33) L7 AR-1260-3	7.839	6.386	357.0E6	151.9E6	1392.140	900.916 #
34) L7 AR-1260-4	8.066	6.850	429.0E6	127.6E6	1504.708	1165.321
35) L7 AR-1260-5	8.388	7.089	959.8E6	318.7E6	1717.087	1246.612 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO080724\
Data File : PO105264.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Aug 2024 17:48
Operator : YP/AJ
Sample : P3491-03 5X
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
COMP-1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 08 02:41:11 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 30 07:11:20 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.433 min
Delta R.T.: 0.004 min
Response: 18149540
Conc: 2.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
COMP-1




#1 Tetrachloro-m-xylene

R.T.: 3.546 min
Delta R.T.: 0.006 min
Response: 6737124
Conc: 2.64 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.220 min
Delta R.T.: -0.001 min
Response: 26289524
Conc: 5.12 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.428 min
Delta R.T.: -0.009 min
Response: 7293908
Conc: 4.59 ng/ml



#31 AR-1260-1


R.T.: 7.218 min
Delta R.T.: 0.003 min
Response: 374094918
Conc: 1178.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
COMP-1




#31 AR-1260-1

R.T.: 6.051 min
Delta R.T.: -0.001 min
Response: 133089739
Conc: 977.18 ng/ml




#32 AR-1260-2

R.T.: 7.477 min
Delta R.T.: 0.003 min
Response: 463224217
Conc: 1273.76 ng/ml



#32 AR-1260-2


R.T.: 6.236 min
Delta R.T.: -0.002 min
Response: 163511162
Conc: 971.77 ng/ml



#33 AR-1260-3


R.T.: 7.839 min
Delta R.T.: 0.002 min
Response: 356954432
Conc: 1392.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
COMP-1




#33 AR-1260-3

R.T.: 6.386 min
Delta R.T.: -0.003 min
Response: 151945028
Conc: 900.92 ng/ml




#34 AR-1260-4

R.T.: 8.066 min
Delta R.T.: 0.002 min
Response: 428960532
Conc: 1504.71 ng/ml



#34 AR-1260-4


R.T.: 6.850 min
Delta R.T.: -0.004 min
Response: 127615092
Conc: 1165.32 ng/ml



#35 AR-1260-5

R.T.: 8.388 min
Delta R.T.: 0.000 min
Response: 959772842
Conc: 1717.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
COMP-1



#35 AR-1260-5

R.T.: 7.089 min
Delta R.T.: -0.005 min
Response: 318655358
Conc: 1246.61 ng/ml