

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0080822\
 Data File : P0088695.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Aug 2022 11:27
 Operator : YP/AJ
 Sample : N4031-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 S2-02-B3

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 08 13:14:40 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 05 23:24:09 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.436	3.594	56410278	21025240	20.913	19.426
2) SA Decachlor...	10.270	8.585	30570739	18304440	15.842	15.104

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO080822\
Data File : PO088695.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Aug 2022 11:27
Operator : YP/AJ
Sample : N4031-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
S2-02-B3

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 08 13:14:40 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO080422.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 05 23:24:09 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.436 min
Delta R.T.: -0.003 min
Response: 56410278
Conc: 20.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
S2-02-B3




#1 Tetrachloro-m-xylene

R.T.: 3.594 min
Delta R.T.: -0.004 min
Response: 21025240
Conc: 19.43 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.270 min
Delta R.T.: -0.007 min
Response: 30570739
Conc: 15.84 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.585 min
Delta R.T.: -0.006 min
Response: 18304440
Conc: 15.10 ng/ml