

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0080823\
 Data File : P0096757.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Aug 2023 12:27
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 09 01:23:47 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 08 04:34:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.409	3.620	170.9E6	68770848	51.640	51.950
2) SA Decachlor...	10.349	8.643	113.4E6	59170073	51.841	51.144
Target Compounds						
21) L5 AR-1248-1	5.594	4.706	30030080	14165334	512.511	504.778m
22) L5 AR-1248-2	5.873	4.945	47387261	20806518	504.025	490.940
23) L5 AR-1248-3	6.081	4.987	49774315	21744619	508.894	495.716
24) L5 AR-1248-4	6.493	5.159	46112458	24912996	507.604	488.013
25) L5 AR-1248-5	6.533	5.552	45011295	22742985	510.993	486.462



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO080823\
Data File : PO096757.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Aug 2023 12:27
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 09 01:23:47 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO080723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 08 04:34:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.409 min
Delta R.T.: 0.002 min
Response: 170936587
Conc: 51.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.620 min
Delta R.T.: 0.000 min
Response: 68770848
Conc: 51.95 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.349 min
Delta R.T.: 0.009 min
Response: 113438851
Conc: 51.84 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.643 min
Delta R.T.: 0.000 min
Response: 59170073
Conc: 51.14 ng/ml



#21 AR-1248-1


R.T.: 5.594 min
Delta R.T.: 0.003 min
Response: 30030080
Conc: 512.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.706 min
Delta R.T.: -0.002 min
Response: 14165334
Conc: 504.78 ng/ml m




#22 AR-1248-2

R.T.: 5.873 min
Delta R.T.: 0.003 min
Response: 47387261
Conc: 504.03 ng/ml



#22 AR-1248-2


R.T.: 4.945 min
Delta R.T.: 0.000 min
Response: 20806518
Conc: 490.94 ng/ml



#23 AR-1248-3


R.T.: 6.081 min
Delta R.T.: 0.004 min
Response: 49774315
Conc: 508.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.987 min
Delta R.T.: 0.000 min
Response: 21744619
Conc: 495.72 ng/ml




#24 AR-1248-4

R.T.: 6.493 min
Delta R.T.: 0.005 min
Response: 46112458
Conc: 507.60 ng/ml



#24 AR-1248-4


R.T.: 5.159 min
Delta R.T.: 0.000 min
Response: 24912996
Conc: 488.01 ng/ml



#25 AR-1248-5

R.T.: 6.533 min
Delta R.T.: 0.006 min
Response: 45011295
Conc: 510.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.552 min
Delta R.T.: -0.001 min
Response: 22742985
Conc: 486.46 ng/ml