

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0080823\
 Data File : P0096788.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Aug 2023 21:41
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 09 01:32:46 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 08 04:34:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.411	3.620	169.4E6	68460767	51.168	51.716
2)	SA Decachlor...	10.350	8.643	115.8E6	59056892	52.916	51.046
Target Compounds							
26)	L6 AR-1254-1	6.468	5.512	53854510	35810338	492.764	485.260
27)	L6 AR-1254-2	6.693	5.660	78263615	31900126	485.423	484.213
28)	L6 AR-1254-3	7.069	6.064	76855223	49125819	490.471	486.093
29)	L6 AR-1254-4	7.362	6.294	50249532	28694899	498.418	495.810
30)	L6 AR-1254-5	7.793	6.713	56723763	42930902	505.495	488.322



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO080823\
Data File : PO096788.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Aug 2023 21:41
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 09 01:32:46 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO080723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 08 04:34:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.411 min
Delta R.T.: 0.004 min
Response: 169374232
Conc: 51.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.620 min
Delta R.T.: 0.000 min
Response: 68460767
Conc: 51.72 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.350 min
Delta R.T.: 0.010 min
Response: 115789310
Conc: 52.92 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.643 min
Delta R.T.: -0.002 min
Response: 59056892
Conc: 51.05 ng/ml



#26 AR-1254-1


R.T.: 6.468 min
Delta R.T.: 0.006 min
Response: 53854510
Conc: 492.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.512 min
Delta R.T.: 0.000 min
Response: 35810338
Conc: 485.26 ng/ml




#27 AR-1254-2

R.T.: 6.693 min
Delta R.T.: 0.007 min
Response: 78263615
Conc: 485.42 ng/ml



#27 AR-1254-2


R.T.: 5.660 min
Delta R.T.: 0.000 min
Response: 31900126
Conc: 484.21 ng/ml



#28 AR-1254-3


R.T.: 7.069 min
Delta R.T.: 0.008 min
Response: 76855223
Conc: 490.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.064 min
Delta R.T.: 0.000 min
Response: 49125819
Conc: 486.09 ng/ml




#29 AR-1254-4

R.T.: 7.362 min
Delta R.T.: 0.008 min
Response: 50249532
Conc: 498.42 ng/ml



#29 AR-1254-4


R.T.: 6.294 min
Delta R.T.: 0.000 min
Response: 28694899
Conc: 495.81 ng/ml



#30 AR-1254-5

R.T.: 7.793 min
Delta R.T.: 0.008 min
Response: 56723763
Conc: 505.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.713 min
Delta R.T.: 0.000 min
Response: 42930902
Conc: 488.32 ng/ml