

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0080823\
 Data File : P0096790.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Aug 2023 22:15
 Operator : YP/AJ
 Sample : PB154674BL
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB154674BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 09 01:33:20 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 08 04:34:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.411	3.620	66331064	25864997	20.038	19.539
2) SA Decachlor...	10.352	8.643	47970137	25628060	21.922	22.152

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO080823\
Data File : PO096790.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Aug 2023 22:15
Operator : YP/AJ
Sample : PB154674BL
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB154674BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 09 01:33:20 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO080723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 08 04:34:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.411 min
Delta R.T.: 0.004 min
Response: 66331064
Conc: 20.04 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB154674BL




#1 Tetrachloro-m-xylene

R.T.: 3.620 min
Delta R.T.: 0.000 min
Response: 25864997
Conc: 19.54 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.352 min
Delta R.T.: 0.012 min
Response: 47970137
Conc: 21.92 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.643 min
Delta R.T.: -0.001 min
Response: 25628060
Conc: 22.15 ng/ml