

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0080918\
Data File : P0047870.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 09 Aug 2018 15:11
Operator : SM/SJ
Sample : J4271-01
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
HR-05-080818-A

Manual Integrations
APPROVED

Sohil
8/10/2018 4:14:07 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 10 07:53:59 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0072718.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jul 27 02:33:29 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.395	3.638	3991150	4587347	19.768m	18.669m
2)	SA Decachlor...	10.089	8.626	2302641	1997129	14.118	12.705m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0080918\
Data File : P0047870.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 09 Aug 2018 15:11
Operator : SM/SJ
Sample : J4271-01
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
HR-05-080818-A


Manual Integrations
APPROVED

Sohil
8/10/2018 4:14:07 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 10 07:53:59 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072718.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jul 27 02:33:29 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.395 min
Delta R.T.: -0.003 min
Response: 3991150
Conc: 19.77 ng/ml

Instrument :
ECD_O
ClientSampleId :
HR-05-080818-A


Manual Integrations
APPROVED

Sohil
8/10/2018 4:14:07 PM




#1 Tetrachloro-m-xylene

R.T.: 3.638 min
Delta R.T.: -0.007 min
Response: 4587347
Conc: 18.67 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.089 min
Delta R.T.: 0.000 min
Response: 2302641
Conc: 14.12 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.626 min
Delta R.T.: -0.018 min
Response: 1997129
Conc: 12.71 ng/ml m