

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0080922\
 Data File : P0088792.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Aug 2022 22:12
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 10 09:20:03 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 05 23:24:09 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.437	3.594	143.0E6	57031568	53.005	52.695
2) SA Decachlor...	10.274	8.586	90243613	56904687	46.766	46.955
Target Compounds						
26) L6 AR-1254-1	6.473	5.471	43269870	35876355	473.687	520.895
27) L6 AR-1254-2	6.693	5.619	65805480	30394854	481.198	513.252
28) L6 AR-1254-3	7.061	6.021	68100215	46569277	483.920	500.808
29) L6 AR-1254-4	7.348	6.250	47122809	28106746	466.354	496.141
30) L6 AR-1254-5	7.769	6.668	46646661	40400178	438.041	494.936



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO080922\
Data File : PO088792.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Aug 2022 22:12
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 10 09:20:03 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO080422.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 05 23:24:09 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.437 min
Delta R.T.: -0.002 min
Response: 142974469
Conc: 53.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.594 min
Delta R.T.: -0.003 min
Response: 57031568
Conc: 52.69 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.274 min
Delta R.T.: -0.003 min
Response: 90243613
Conc: 46.77 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.586 min
Delta R.T.: -0.006 min
Response: 56904687
Conc: 46.95 ng/ml



#26 AR-1254-1


R.T.: 6.473 min
Delta R.T.: -0.002 min
Response: 43269870
Conc: 473.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.471 min
Delta R.T.: -0.003 min
Response: 35876355
Conc: 520.90 ng/ml




#27 AR-1254-2

R.T.: 6.693 min
Delta R.T.: 0.000 min
Response: 65805480
Conc: 481.20 ng/ml



#27 AR-1254-2


R.T.: 5.619 min
Delta R.T.: -0.003 min
Response: 30394854
Conc: 513.25 ng/ml



#28 AR-1254-3


R.T.: 7.061 min
Delta R.T.: -0.001 min
Response: 68100215
Conc: 483.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.021 min
Delta R.T.: -0.004 min
Response: 46569277
Conc: 500.81 ng/ml




#29 AR-1254-4

R.T.: 7.348 min
Delta R.T.: 0.000 min
Response: 47122809
Conc: 466.35 ng/ml



#29 AR-1254-4


R.T.: 6.250 min
Delta R.T.: -0.003 min
Response: 28106746
Conc: 496.14 ng/ml



#30 AR-1254-5

R.T.: 7.769 min
Delta R.T.: 0.000 min
Response: 46646661
Conc: 438.04 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.668 min
Delta R.T.: -0.003 min
Response: 40400178
Conc: 494.94 ng/ml