

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0080922\
 Data File : P0088798.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Aug 2022 23:57
 Operator : YP/AJ
 Sample : N4106-17 10X
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 HIH07719-END-1-2

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/10/2022
 Supervised By :Ankita Jodhani 08/10/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 10 03:05:21 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0080422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 05 23:24:09 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.436	3.594	4673222	2012160	1.733	1.859
2) SA Decachlor...	10.277	8.587	3695877	2038068	1.915	1.682
Target Compounds						
26) L6 AR-1254-1	6.473	5.471	1060129	604238	11.606	8.773
27) L6 AR-1254-2	6.693	5.620	1514000	690851	11.071	11.666
28) L6 AR-1254-3	7.062	6.022	2040270	1016777	14.498	10.934
29) L6 AR-1254-4	7.347	6.250	1967059	1101861	19.467	19.450m
30) L6 AR-1254-5	7.768	6.668	2543507	1588144	23.885	19.456

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO080922\
Data File : PO088798.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Aug 2022 23:57
Operator : YP/AJ
Sample : N4106-17 10X
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
HIH07719-END-1-2


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/10/2022
Supervised By :Ankita Jodhani 08/10/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 10 03:05:21 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO080422.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 05 23:24:09 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 4.436 min
Delta R.T.: -0.003 min
Response: 4673222
Conc: 1.73 ng/ml

Instrument :

ECD_O


ClientSampleId :

HIH07719-END-1-2

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 08/10/2022
Supervised By :Ankita Jodhani 08/10/2022




#1 Tetrachloro-m-xylene

R.T.: 3.594 min
Delta R.T.: -0.003 min
Response: 2012160
Conc: 1.86 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.277 min
Delta R.T.: 0.000 min
Response: 3695877
Conc: 1.92 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.587 min
Delta R.T.: -0.005 min
Response: 2038068
Conc: 1.68 ng/ml




#26 AR-1254-1

R.T.: 6.473 min
Delta R.T.: -0.001 min
Response: 1060129
Conc: 11.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
HIH07719-END-1-2


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/10/2022
Supervised By :Ankita Jodhani 08/10/2022




#26 AR-1254-1

R.T.: 5.471 min
Delta R.T.: -0.003 min
Response: 604238
Conc: 8.77 ng/ml




#27 AR-1254-2

R.T.: 6.693 min
Delta R.T.: 0.000 min
Response: 1514000
Conc: 11.07 ng/ml



#27 AR-1254-2

R.T.: 5.620 min
Delta R.T.: -0.002 min
Response: 690851
Conc: 11.67 ng/ml




#28 AR-1254-3

R.T.: 7.062 min
Delta R.T.: 0.000 min
Response: 2040270
Conc: 14.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
HIH07719-END-1-2


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/10/2022
Supervised By :Ankita Jodhani 08/10/2022




#28 AR-1254-3

R.T.: 6.022 min
Delta R.T.: -0.004 min
Response: 1016777
Conc: 10.93 ng/ml




#29 AR-1254-4

R.T.: 7.347 min
Delta R.T.: -0.001 min
Response: 1967059
Conc: 19.47 ng/ml



#29 AR-1254-4

R.T.: 6.250 min
Delta R.T.: -0.003 min
Response: 1101861
Conc: 19.45 ng/ml m




#30 AR-1254-5

R.T.: 7.768 min
Delta R.T.: -0.001 min
Response: 2543507
Conc: 23.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
HIH07719-END-1-2

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/10/2022
Supervised By :Ankita Jodhani 08/10/2022



#30 AR-1254-5

R.T.: 6.668 min
Delta R.T.: -0.002 min
Response: 1588144
Conc: 19.46 ng/ml