

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0080923\
 Data File : P0096873.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Aug 2023 00:31
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 10 01:39:38 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 08 04:34:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.411	3.619	164.8E6	67651320	49.775	51.104
2) SA Decachlor...	10.352	8.639	113.1E6	55304742	51.694	47.803
Target Compounds						
26) L6 AR-1254-1	6.470	5.508	50862617	35397676	465.388	479.668
27) L6 AR-1254-2	6.694	5.657	75404786	30816688	467.692	467.768
28) L6 AR-1254-3	7.071	6.061	75087889	48424303	479.192	479.152
29) L6 AR-1254-4	7.364	6.291	50529981	28666831	501.199	495.325
30) L6 AR-1254-5	7.795	6.710	56698130	42615133	505.266	484.731



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO080923\
Data File : PO096873.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Aug 2023 00:31
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 10 01:39:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO080723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 08 04:34:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.411 min
Delta R.T.: 0.004 min
Response: 164765072
Conc: 49.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.619 min
Delta R.T.: -0.003 min
Response: 67651320
Conc: 51.10 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.352 min
Delta R.T.: 0.013 min
Response: 113115417
Conc: 51.69 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.639 min
Delta R.T.: -0.005 min
Response: 55304742
Conc: 47.80 ng/ml



#26 AR-1254-1


R.T.: 6.470 min
Delta R.T.: 0.008 min
Response: 50862617
Conc: 465.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.508 min
Delta R.T.: -0.003 min
Response: 35397676
Conc: 479.67 ng/ml




#27 AR-1254-2

R.T.: 6.694 min
Delta R.T.: 0.008 min
Response: 75404786
Conc: 467.69 ng/ml



#27 AR-1254-2


R.T.: 5.657 min
Delta R.T.: -0.003 min
Response: 30816688
Conc: 467.77 ng/ml



#28 AR-1254-3


R.T.: 7.071 min
Delta R.T.: 0.010 min
Response: 75087889
Conc: 479.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.061 min
Delta R.T.: -0.004 min
Response: 48424303
Conc: 479.15 ng/ml




#29 AR-1254-4

R.T.: 7.364 min
Delta R.T.: 0.010 min
Response: 50529981
Conc: 501.20 ng/ml



#29 AR-1254-4


R.T.: 6.291 min
Delta R.T.: -0.003 min
Response: 28666831
Conc: 495.32 ng/ml



#30 AR-1254-5

R.T.: 7.795 min
Delta R.T.: 0.010 min
Response: 56698130
Conc: 505.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.710 min
Delta R.T.: -0.003 min
Response: 42615133
Conc: 484.73 ng/ml