

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081022\
 Data File : P0088851.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Aug 2022 20:53
 Operator : YP/AJ
 Sample : N4123-05 10X
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 LAW-2022-36A

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 10 22:00:09 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 05 23:24:09 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.435	3.594	4165528	1601398	1.544	1.480m
2) SA Decachlor...	10.272	8.584	3109357	1809246	1.611	1.493m

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081022\
Data File : PO088851.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Aug 2022 20:53
Operator : YP/AJ
Sample : N4123-05 10X
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
LAW-2022-36A

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 10 22:00:09 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO080422.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 05 23:24:09 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.435 min
Delta R.T.: -0.004 min
Response: 4165528
Conc: 1.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
LAW-2022-36A




#1 Tetrachloro-m-xylene

R.T.: 3.594 min
Delta R.T.: -0.004 min
Response: 1601398
Conc: 1.48 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.272 min
Delta R.T.: -0.005 min
Response: 3109357
Conc: 1.61 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.584 min
Delta R.T.: -0.007 min
Response: 1809246
Conc: 1.49 ng/ml m