

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081022\
 Data File : P0088874.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Aug 2022 06:34
 Operator : YP/AJ
 Sample : N4123-14 10X
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 LAW-2022-41A

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 11 08:39:36 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 05 23:24:09 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.434	3.593	4320425	1595818	1.602m	1.474
2) SA Decachlor...	10.270	8.582	3693928	1810753	1.914	1.494

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081022\
Data File : PO088874.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Aug 2022 06:34
Operator : YP/AJ
Sample : N4123-14 10X
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
LAW-2022-41A

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 11 08:39:36 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO080422.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 05 23:24:09 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.434 min
Delta R.T.: -0.005 min
Response: 4320425
Conc: 1.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
LAW-2022-41A




#1 Tetrachloro-m-xylene

R.T.: 3.593 min
Delta R.T.: -0.004 min
Response: 1595818
Conc: 1.47 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.270 min
Delta R.T.: -0.007 min
Response: 3693928
Conc: 1.91 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.582 min
Delta R.T.: -0.010 min
Response: 1810753
Conc: 1.49 ng/ml