

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081023\
 Data File : P0096923.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Aug 2023 15:35
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 10 21:44:55 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 08 04:34:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.413	3.619	166.4E6	68641743	50.280	51.852
2) SA Decachlor...	10.356	8.641	112.0E6	58376211	51.189	50.458
Target Compounds						
21) L5 AR-1248-1	5.599	4.705	30077314	14078841	513.317	501.696
22) L5 AR-1248-2	5.878	4.943	46399143	20785450	493.516	490.443
23) L5 AR-1248-3	6.087	4.985	49110507	21932324	502.107	499.996
24) L5 AR-1248-4	6.500	5.157	46438758	25481335	511.195	499.146
25) L5 AR-1248-5	6.539	5.551	45022607	23533686	511.121	503.374



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081023\
Data File : P0096923.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Aug 2023 15:35
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 10 21:44:55 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 08 04:34:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.413 min
Delta R.T.: 0.006 min
Response: 166434995
Conc: 50.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.619 min
Delta R.T.: -0.002 min
Response: 68641743
Conc: 51.85 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.356 min
Delta R.T.: 0.017 min
Response: 112010841
Conc: 51.19 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.641 min
Delta R.T.: -0.004 min
Response: 58376211
Conc: 50.46 ng/ml



#21 AR-1248-1


R.T.: 5.599 min
Delta R.T.: 0.008 min
Response: 30077314
Conc: 513.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.705 min
Delta R.T.: -0.003 min
Response: 14078841
Conc: 501.70 ng/ml




#22 AR-1248-2

R.T.: 5.878 min
Delta R.T.: 0.008 min
Response: 46399143
Conc: 493.52 ng/ml



#22 AR-1248-2


R.T.: 4.943 min
Delta R.T.: -0.003 min
Response: 20785450
Conc: 490.44 ng/ml



#23 AR-1248-3


R.T.: 6.087 min
Delta R.T.: 0.010 min
Response: 49110507
Conc: 502.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.985 min
Delta R.T.: -0.003 min
Response: 21932324
Conc: 500.00 ng/ml




#24 AR-1248-4

R.T.: 6.500 min
Delta R.T.: 0.012 min
Response: 46438758
Conc: 511.20 ng/ml



#24 AR-1248-4


R.T.: 5.157 min
Delta R.T.: -0.002 min
Response: 25481335
Conc: 499.15 ng/ml



#25 AR-1248-5

R.T.: 6.539 min
Delta R.T.: 0.012 min
Response: 45022607
Conc: 511.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.551 min
Delta R.T.: -0.003 min
Response: 23533686
Conc: 503.37 ng/ml