

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081024\
 Data File : P0105383.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Aug 2024 23:16
 Operator : YP/AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 10 00:32:28 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 10 00:29:34 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.429	3.537	657.1E6	241.5E6	74.472	74.535
2)	SA Decachlor...	10.220	8.428	437.4E6	131.8E6	74.719	73.926
Target Compounds							
26)	L6 AR-1254-1	6.461	5.375	222.1E6	118.6E6	741.282	744.205
27)	L6 AR-1254-2	6.681	5.521	329.4E6	105.0E6	744.377	741.832
28)	L6 AR-1254-3	7.051	5.917	341.1E6	165.0E6	742.934	743.261
29)	L6 AR-1254-4	7.338	6.142	258.0E6	103.4E6	743.032	743.991
30)	L6 AR-1254-5	7.761	6.554	287.3E6	146.3E6	742.416	746.935



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081024\
Data File : PO105383.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Aug 2024 23:16
Operator : YP/AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 10 00:32:28 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO081024.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 10 00:29:34 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.429 min
Delta R.T.: 0.000 min
Response: 657061481
Conc: 74.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.537 min
Delta R.T.: 0.000 min
Response: 241547246
Conc: 74.53 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.220 min
Delta R.T.: -0.001 min
Response: 437420201
Conc: 74.72 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.428 min
Delta R.T.: 0.000 min
Response: 131808756
Conc: 73.93 ng/ml



#26 AR-1254-1


R.T.: 6.461 min
Delta R.T.: 0.000 min
Response: 222059853
Conc: 741.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC750




#26 AR-1254-1

R.T.: 5.375 min
Delta R.T.: 0.000 min
Response: 118554453
Conc: 744.21 ng/ml




#27 AR-1254-2

R.T.: 6.681 min
Delta R.T.: 0.000 min
Response: 329431499
Conc: 744.38 ng/ml



#27 AR-1254-2


R.T.: 5.521 min
Delta R.T.: 0.000 min
Response: 104993620
Conc: 741.83 ng/ml



#28 AR-1254-3


R.T.: 7.051 min
Delta R.T.: 0.000 min
Response: 341099913
Conc: 742.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC750




#28 AR-1254-3

R.T.: 5.917 min
Delta R.T.: 0.000 min
Response: 165018012
Conc: 743.26 ng/ml




#29 AR-1254-4

R.T.: 7.338 min
Delta R.T.: 0.000 min
Response: 257990376
Conc: 743.03 ng/ml



#29 AR-1254-4


R.T.: 6.142 min
Delta R.T.: 0.000 min
Response: 103383112
Conc: 743.99 ng/ml



#30 AR-1254-5

R.T.: 7.761 min
Delta R.T.: 0.000 min
Response: 287341715
Conc: 742.42 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC750



#30 AR-1254-5

R.T.: 6.554 min
Delta R.T.: 0.000 min
Response: 146257916
Conc: 746.93 ng/ml