

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081122\
 Data File : P0088880.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Aug 2022 09:49
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 11 12:44:47 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 05 23:24:09 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.435	3.592	140.2E6	53800153	51.966	49.709
2) SA Decachlor...	10.279	8.585	103.1E6	57618865	53.405	47.544
Target Compounds						
21) L5 AR-1248-1	5.612	4.670	27716053	12304635	537.218	524.029
22) L5 AR-1248-2	5.888	4.905	39585891	18611703	515.656	518.590
23) L5 AR-1248-3	6.093	4.946	42469694	19655298	514.166	516.212
24) L5 AR-1248-4	6.500	5.118	47433149	24692596	528.164	537.701
25) L5 AR-1248-5	6.539	5.511	48032068	21616023	541.869	495.090



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081122\
Data File : PO088880.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Aug 2022 09:49
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 11 12:44:47 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO080422.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 05 23:24:09 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.435 min
Delta R.T.: -0.004 min
Response: 140170464
Conc: 51.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.592 min
Delta R.T.: -0.005 min
Response: 53800153
Conc: 49.71 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.279 min
Delta R.T.: 0.002 min
Response: 103055092
Conc: 53.41 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.585 min
Delta R.T.: -0.007 min
Response: 57618865
Conc: 47.54 ng/ml



#21 AR-1248-1


R.T.: 5.612 min
Delta R.T.: -0.002 min
Response: 27716053
Conc: 537.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.670 min
Delta R.T.: -0.005 min
Response: 12304635
Conc: 524.03 ng/ml




#22 AR-1248-2

R.T.: 5.888 min
Delta R.T.: -0.002 min
Response: 39585891
Conc: 515.66 ng/ml



#22 AR-1248-2


R.T.: 4.905 min
Delta R.T.: -0.006 min
Response: 18611703
Conc: 518.59 ng/ml



#23 AR-1248-3


R.T.: 6.093 min
Delta R.T.: -0.002 min
Response: 42469694
Conc: 514.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.946 min
Delta R.T.: -0.006 min
Response: 19655298
Conc: 516.21 ng/ml




#24 AR-1248-4

R.T.: 6.500 min
Delta R.T.: -0.002 min
Response: 47433149
Conc: 528.16 ng/ml



#24 AR-1248-4


R.T.: 5.118 min
Delta R.T.: -0.006 min
Response: 24692596
Conc: 537.70 ng/ml



#25 AR-1248-5

R.T.: 6.539 min
Delta R.T.: -0.001 min
Response: 48032068
Conc: 541.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.511 min
Delta R.T.: -0.005 min
Response: 21616023
Conc: 495.09 ng/ml