

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081123\
 Data File : P0097008.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Aug 2023 03:22
 Operator : YP/AJ
 Sample : 03966-05 10X
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 KMA576Q-END-1-1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 12 04:09:02 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 08 04:34:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.410	3.616	4272670	1853388	1.291	1.400
2) SA Decachlor...	10.349	8.633	3386817	1941730	1.548m	1.678m

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081123\
Data File : P0097008.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2023 03:22
Operator : YP/AJ
Sample : 03966-05 10X
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
KMA576Q-END-1-1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 12 04:09:02 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 08 04:34:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.410 min
Delta R.T.: 0.003 min
Response: 4272670
Conc: 1.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
KMA576Q-END-1-1




#1 Tetrachloro-m-xylene

R.T.: 3.616 min
Delta R.T.: -0.005 min
Response: 1853388
Conc: 1.40 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.349 min
Delta R.T.: 0.009 min
Response: 3386817
Conc: 1.55 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.633 min
Delta R.T.: -0.011 min
Response: 1941730
Conc: 1.68 ng/ml m