

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0081224\
 Data File : P0105455.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Aug 2024 01:54
 Operator : YP/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 13 03:21:58 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0081024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 10 04:08:56 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.426	3.535	175.8E6	68128057	20.130	21.690
2)	SA Decachlor...	10.216	8.426	133.9E6	42876678	23.278	24.219
Target Compounds							
3)	L1 AR-1016-1	5.591	4.602	66527	661347	0.220	5.885 #
4)	L1 AR-1016-2	5.628	4.602	426717	661347	0.998	4.248 #
5)	L1 AR-1016-3	5.668	4.787	1621028	197800	6.318	2.362 #
6)	L1 AR-1016-4	5.789	4.832	1381963	419291	6.319	6.227
7)	L1 AR-1016-5	6.078	5.027	290078	415551	1.356	4.965 #
8)	L2 AR-1221-1	4.636	3.737	300316	2825	2.678	0.072 #
9)	L2 AR-1221-2	4.733	3.826	3139072	1242804	37.035	40.453
10)	L2 AR-1221-3	4.810	3.899	304522	49674	1.252	0.555 #
11)	L3 AR-1232-1	4.810	3.899	304522	49674	1.523	0.680 #
12)	L3 AR-1232-2	5.333	4.602	223898	661347	2.115	9.419 #
13)	L3 AR-1232-3	5.628	4.787	426717	197800	2.229	5.176 #
14)	L3 AR-1232-4	5.789	4.864	1381963	137033	14.169	3.939 #
15)	L3 AR-1232-5	5.865	5.027	590993	415551	8.201	11.772 #
16)	L4 AR-1242-1	5.591	4.602	66527	661347	0.286	7.498 #
17)	L4 AR-1242-2	5.628	4.602	426717	661347	1.309	5.570 #
18)	L4 AR-1242-3	5.668	4.787	1621028	197800	8.217	3.046 #
19)	L4 AR-1242-4	5.789	4.864	1381963	137033	8.100	2.066 #
20)	L4 AR-1242-5	6.527	5.371	1266002	607783	7.244	7.396
21)	L5 AR-1248-1	5.591	4.602	66527	661347	0.364	9.469 #
22)	L5 AR-1248-2	5.865	4.832	590993	419291	2.346	4.418 #
23)	L5 AR-1248-3	6.078	4.864	290078	137033	1.026	1.373 #
24)	L5 AR-1248-4	6.492	5.027	2362489	415551	7.389	3.682 #
25)	L5 AR-1248-5	6.527	5.397	1266002	2794635	4.093	24.818 #
26)	L6 AR-1254-1	6.459	5.371	6753195	607783	21.535	3.733 #
27)	L6 AR-1254-2	6.657	5.528	906142	772404	2.008	5.280 #
28)	L6 AR-1254-3	7.051	5.917	6557989	440579	13.824	1.953 #
29)	L6 AR-1254-4	7.339	6.155	866715	842865	2.420	5.973 #
30)	L6 AR-1254-5	7.766	6.549	832260	311619	2.097	1.555 #
31)	L7 AR-1260-1	7.231	6.046	3078791	598911	8.400	3.837 #
32)	L7 AR-1260-2	7.466	6.215	1707027	1187563	4.100	6.283 #
33)	L7 AR-1260-3	7.833	6.386	3521056	3593915	13.604	19.454 #
34)	L7 AR-1260-4	8.081	6.832	8000647	46154	24.250	0.378 #
35)	L7 AR-1260-5	8.402	7.084	5634149	339529	9.471	1.220 #
36)	L8 AR-1262-1	7.766	6.549	832260	311619	2.620	2.929
37)	L8 AR-1262-2	8.402	6.832	5634149	46154	7.490	0.272 #
38)	L8 AR-1262-3	8.706	7.374	746449	543420	1.517	4.304 #
39)	L8 AR-1262-4	8.793	7.426	158790	383374	0.432	1.656 #
40)	L8 AR-1262-5	9.439	7.911	35202	51644	0.131	0.514 #
41)	L9 AR-1268-1	8.706	7.374	746449	543420	0.918	1.557 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081224\
Data File : P0105455.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Aug 2024 01:54
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 13 03:21:58 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081024.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 10 04:08:56 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.793	7.426	158790	383374	0.218	1.229 #
43)	L9 AR-1268-3	9.028	7.633	434327	488878	0.673	1.934 #
44)	L9 AR-1268-4	9.439	7.911	35202	51644	0.126	0.480 #
45)	L9 AR-1268-5	9.874	8.191	1521456	441893	0.728	0.661



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081224\
Data File : PO105455.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Aug 2024 01:54
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 13 03:21:58 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO081024.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 10 04:08:56 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.426 min
Delta R.T.: -0.003 min
Response: 175823913
Conc: 20.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.535 min
Delta R.T.: -0.002 min
Response: 68128057
Conc: 21.69 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.216 min
Delta R.T.: -0.006 min
Response: 133906776
Conc: 23.28 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.426 min
Delta R.T.: -0.002 min
Response: 42876678
Conc: 24.22 ng/ml



#3 AR-1016-1


R.T.: 5.591 min
Delta R.T.: -0.010 min
Response: 66527
Conc: 0.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#3 AR-1016-1

R.T.: 4.602 min
Delta R.T.: 0.010 min
Response: 661347
Conc: 5.88 ng/ml




#4 AR-1016-2

R.T.: 5.628 min
Delta R.T.: 0.005 min
Response: 426717
Conc: 1.00 ng/ml



#4 AR-1016-2


R.T.: 4.602 min
Delta R.T.: -0.008 min
Response: 661347
Conc: 4.25 ng/ml



#5 AR-1016-3


R.T.: 5.668 min
Delta R.T.: -0.017 min
Response: 1621028
Conc: 6.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#5 AR-1016-3

R.T.: 4.787 min
Delta R.T.: 0.005 min
Response: 197800
Conc: 2.36 ng/ml




#6 AR-1016-4

R.T.: 5.789 min
Delta R.T.: 0.004 min
Response: 1381963
Conc: 6.32 ng/ml



#6 AR-1016-4


R.T.: 4.832 min
Delta R.T.: 0.008 min
Response: 419291
Conc: 6.23 ng/ml



#7 AR-1016-5


R.T.: 6.078 min
Delta R.T.: -0.004 min
Response: 290078
Conc: 1.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#7 AR-1016-5

R.T.: 5.027 min
Delta R.T.: -0.004 min
Response: 415551
Conc: 4.97 ng/ml




#8 AR-1221-1

R.T.: 4.636 min
Delta R.T.: 0.000 min
Response: 300316
Conc: 2.68 ng/ml



#8 AR-1221-1


R.T.: 3.737 min
Delta R.T.: -0.007 min
Response: 2825
Conc: 0.07 ng/ml



#9 AR-1221-2


R.T.: 4.733 min
Delta R.T.: 0.010 min
Response: 3139072
Conc: 37.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#9 AR-1221-2

R.T.: 3.826 min
Delta R.T.: -0.003 min
Response: 1242804
Conc: 40.45 ng/ml




#10 AR-1221-3

R.T.: 4.810 min
Delta R.T.: 0.011 min
Response: 304522
Conc: 1.25 ng/ml



#10 AR-1221-3


R.T.: 3.899 min
Delta R.T.: -0.003 min
Response: 49674
Conc: 0.56 ng/ml



#11 AR-1232-1


R.T.: 4.810 min
Delta R.T.: 0.011 min
Response: 304522
Conc: 1.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#11 AR-1232-1

R.T.: 3.899 min
Delta R.T.: -0.002 min
Response: 49674
Conc: 0.68 ng/ml




#12 AR-1232-2

R.T.: 5.333 min
Delta R.T.: 0.002 min
Response: 223898
Conc: 2.12 ng/ml



#12 AR-1232-2


R.T.: 4.602 min
Delta R.T.: -0.008 min
Response: 661347
Conc: 9.42 ng/ml



#13 AR-1232-3


R.T.: 5.628 min
Delta R.T.: 0.005 min
Response: 426717
Conc: 2.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#13 AR-1232-3

R.T.: 4.787 min
Delta R.T.: 0.005 min
Response: 197800
Conc: 5.18 ng/ml




#14 AR-1232-4

R.T.: 5.789 min
Delta R.T.: 0.004 min
Response: 1381963
Conc: 14.17 ng/ml



#14 AR-1232-4


R.T.: 4.864 min
Delta R.T.: 0.000 min
Response: 137033
Conc: 3.94 ng/ml



#15 AR-1232-5


R.T.: 5.865 min
Delta R.T.: -0.010 min
Response: 590993
Conc: 8.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#15 AR-1232-5

R.T.: 5.027 min
Delta R.T.: -0.004 min
Response: 415551
Conc: 11.77 ng/ml




#16 AR-1242-1

R.T.: 5.591 min
Delta R.T.: -0.010 min
Response: 66527
Conc: 0.29 ng/ml



#16 AR-1242-1


R.T.: 4.602 min
Delta R.T.: 0.010 min
Response: 661347
Conc: 7.50 ng/ml



#17 AR-1242-2


R.T.: 5.628 min
Delta R.T.: 0.005 min
Response: 426717
Conc: 1.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#17 AR-1242-2

R.T.: 4.602 min
Delta R.T.: -0.008 min
Response: 661347
Conc: 5.57 ng/ml




#18 AR-1242-3

R.T.: 5.668 min
Delta R.T.: -0.018 min
Response: 1621028
Conc: 8.22 ng/ml



#18 AR-1242-3


R.T.: 4.787 min
Delta R.T.: 0.006 min
Response: 197800
Conc: 3.05 ng/ml



#19 AR-1242-4


R.T.: 5.789 min
Delta R.T.: 0.004 min
Response: 1381963
Conc: 8.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#19 AR-1242-4

R.T.: 4.864 min
Delta R.T.: 0.000 min
Response: 137033
Conc: 2.07 ng/ml




#20 AR-1242-5

R.T.: 6.527 min
Delta R.T.: -0.001 min
Response: 1266002
Conc: 7.24 ng/ml



#20 AR-1242-5


R.T.: 5.371 min
Delta R.T.: -0.005 min
Response: 607783
Conc: 7.40 ng/ml



#21 AR-1248-1


R.T.: 5.591 min
Delta R.T.: -0.009 min
Response: 66527
Conc: 0.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#21 AR-1248-1

R.T.: 4.602 min
Delta R.T.: 0.011 min
Response: 661347
Conc: 9.47 ng/ml




#22 AR-1248-2

R.T.: 5.865 min
Delta R.T.: -0.010 min
Response: 590993
Conc: 2.35 ng/ml



#22 AR-1248-2


R.T.: 4.832 min
Delta R.T.: 0.010 min
Response: 419291
Conc: 4.42 ng/ml



#23 AR-1248-3


R.T.: 6.078 min
Delta R.T.: -0.003 min
Response: 290078
Conc: 1.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#23 AR-1248-3

R.T.: 4.864 min
Delta R.T.: 0.001 min
Response: 137033
Conc: 1.37 ng/ml




#24 AR-1248-4

R.T.: 6.492 min
Delta R.T.: 0.005 min
Response: 2362489
Conc: 7.39 ng/ml



#24 AR-1248-4


R.T.: 5.027 min
Delta R.T.: -0.003 min
Response: 415551
Conc: 3.68 ng/ml



#25 AR-1248-5


R.T.: 6.527 min
Delta R.T.: 0.000 min
Response: 1266002
Conc: 4.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#25 AR-1248-5

R.T.: 5.397 min
Delta R.T.: -0.018 min
Response: 2794635
Conc: 24.82 ng/ml




#26 AR-1254-1

R.T.: 6.459 min
Delta R.T.: -0.002 min
Response: 6753195
Conc: 21.53 ng/ml



#26 AR-1254-1


R.T.: 5.371 min
Delta R.T.: -0.005 min
Response: 607783
Conc: 3.73 ng/ml



#27 AR-1254-2


R.T.: 6.657 min
Delta R.T.: -0.024 min
Response: 906142
Conc: 2.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#27 AR-1254-2

R.T.: 5.528 min
Delta R.T.: 0.007 min
Response: 772404
Conc: 5.28 ng/ml




#28 AR-1254-3

R.T.: 7.051 min
Delta R.T.: 0.000 min
Response: 6557989
Conc: 13.82 ng/ml



#28 AR-1254-3


R.T.: 5.917 min
Delta R.T.: 0.000 min
Response: 440579
Conc: 1.95 ng/ml



#29 AR-1254-4


R.T.: 7.339 min
Delta R.T.: 0.001 min
Response: 866715
Conc: 2.42 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#29 AR-1254-4

R.T.: 6.155 min
Delta R.T.: 0.013 min
Response: 842865
Conc: 5.97 ng/ml




#30 AR-1254-5

R.T.: 7.766 min
Delta R.T.: 0.004 min
Response: 832260
Conc: 2.10 ng/ml



#30 AR-1254-5


R.T.: 6.549 min
Delta R.T.: -0.005 min
Response: 311619
Conc: 1.56 ng/ml



#31 AR-1260-1


R.T.: 7.231 min
Delta R.T.: 0.014 min
Response: 3078791
Conc: 8.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#31 AR-1260-1

R.T.: 6.046 min
Delta R.T.: 0.000 min
Response: 598911
Conc: 3.84 ng/ml




#32 AR-1260-2

R.T.: 7.466 min
Delta R.T.: -0.009 min
Response: 1707027
Conc: 4.10 ng/ml



#32 AR-1260-2


R.T.: 6.215 min
Delta R.T.: -0.018 min
Response: 1187563
Conc: 6.28 ng/ml



#33 AR-1260-3


R.T.: 7.833 min
Delta R.T.: -0.005 min
Response: 3521056
Conc: 13.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#33 AR-1260-3

R.T.: 6.386 min
Delta R.T.: 0.004 min
Response: 3593915
Conc: 19.45 ng/ml




#34 AR-1260-4

R.T.: 8.081 min
Delta R.T.: 0.015 min
Response: 8000647
Conc: 24.25 ng/ml



#34 AR-1260-4


R.T.: 6.832 min
Delta R.T.: -0.017 min
Response: 46154
Conc: 0.38 ng/ml



#35 AR-1260-5


R.T.: 8.402 min
Delta R.T.: 0.013 min
Response: 5634149
Conc: 9.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#35 AR-1260-5

R.T.: 7.084 min
Delta R.T.: -0.004 min
Response: 339529
Conc: 1.22 ng/ml




#36 AR-1262-1

R.T.: 7.766 min
Delta R.T.: 0.007 min
Response: 832260
Conc: 2.62 ng/ml



#36 AR-1262-1


R.T.: 6.549 min
Delta R.T.: -0.004 min
Response: 311619
Conc: 2.93 ng/ml



#37 AR-1262-2


R.T.: 8.402 min
Delta R.T.: 0.014 min
Response: 5634149
Conc: 7.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#37 AR-1262-2

R.T.: 6.832 min
Delta R.T.: -0.016 min
Response: 46154
Conc: 0.27 ng/ml




#38 AR-1262-3

R.T.: 8.706 min
Delta R.T.: 0.000 min
Response: 746449
Conc: 1.52 ng/ml



#38 AR-1262-3


R.T.: 7.374 min
Delta R.T.: 0.007 min
Response: 543420
Conc: 4.30 ng/ml



#39 AR-1262-4


R.T.: 8.793 min
Delta R.T.: -0.002 min
Response: 158790
Conc: 0.43 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#39 AR-1262-4

R.T.: 7.426 min
Delta R.T.: -0.004 min
Response: 383374
Conc: 1.66 ng/ml




#40 AR-1262-5

R.T.: 9.439 min
Delta R.T.: -0.020 min
Response: 35202
Conc: 0.13 ng/ml



#40 AR-1262-5


R.T.: 7.911 min
Delta R.T.: -0.008 min
Response: 51644
Conc: 0.51 ng/ml



#41 AR-1268-1


R.T.: 8.706 min
Delta R.T.: 0.004 min
Response: 746449
Conc: 0.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#41 AR-1268-1

R.T.: 7.374 min
Delta R.T.: 0.006 min
Response: 543420
Conc: 1.56 ng/ml




#42 AR-1268-2

R.T.: 8.793 min
Delta R.T.: -0.005 min
Response: 158790
Conc: 0.22 ng/ml



#42 AR-1268-2


R.T.: 7.426 min
Delta R.T.: -0.006 min
Response: 383374
Conc: 1.23 ng/ml



#43 AR-1268-3


R.T.: 9.028 min
Delta R.T.: -0.001 min
Response: 434327
Conc: 0.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#43 AR-1268-3

R.T.: 7.633 min
Delta R.T.: 0.000 min
Response: 488878
Conc: 1.93 ng/ml




#44 AR-1268-4

R.T.: 9.439 min
Delta R.T.: -0.020 min
Response: 35202
Conc: 0.13 ng/ml



#44 AR-1268-4


R.T.: 7.911 min
Delta R.T.: -0.008 min
Response: 51644
Conc: 0.48 ng/ml



#45 AR-1268-5

R.T.: 9.874 min
Delta R.T.: -0.005 min
Response: 1521456
Conc: 0.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK



#45 AR-1268-5

R.T.: 8.191 min
Delta R.T.: -0.002 min
Response: 441893
Conc: 0.66 ng/ml