

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081225\
 Data File : P0112854.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Aug 2025 10:05
 Operator : YP/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 12 11:05:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 24 04:54:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.664	3.658	170.0E6	101.6E6	20.911	20.456
2)	SA Decachlor...	8.685	8.632	126.0E6	36010722	17.231	20.378
Target Compounds							
3)	L1 AR-1016-1	4.818f	4.737	207474	147142	0.773	0.841
4)	L1 AR-1016-2	4.818f	4.748	207474	236047	0.511	0.906 #
5)	L1 AR-1016-3	4.818	4.932	207474	126024	0.793	0.927
6)	L1 AR-1016-4	4.945	4.975	58929	128539	0.278	1.157 #
7)	L1 AR-1016-5	5.198	5.183	1515050	231546	6.903	1.616 #
8)	L2 AR-1221-1	3.867	3.875	15226	156746	0.158	2.330 #
9)	L2 AR-1221-2	3.967	3.957	2377395	1699978	33.083	33.512
10)	L2 AR-1221-3	4.062	4.034	258883	46668	1.089	0.294 #
11)	L3 AR-1232-1	4.062	4.034	258883	46668	1.385	0.369 #
12)	L3 AR-1232-2	4.531	4.748	707238	236047	6.571	1.800 #
13)	L3 AR-1232-3	4.818f	4.932	207474	126024	1.003	1.875 #
14)	L3 AR-1232-4	4.945	5.019	58929	83611	0.563	1.397 #
15)	L3 AR-1232-5	4.995	5.183	247652	231546	3.648	3.551
16)	L4 AR-1242-1	4.818f	4.737	207474	147142	0.838	0.929
17)	L4 AR-1242-2	4.818f	4.748	207474	236047	0.561	1.010 #
18)	L4 AR-1242-3	4.818	4.932	207474	126024	0.877	1.040
19)	L4 AR-1242-4	4.945	5.019	58929	83611	0.303	0.681 #
20)	L4 AR-1242-5	5.643f	5.533	13646084	423661	63.792	2.632 #
21)	L5 AR-1248-1	4.818f	4.737	207474	147142	1.108	1.211
22)	L5 AR-1248-2	4.995	4.975	247652	128539	0.971	0.759
23)	L5 AR-1248-3	5.198	5.019	1515050	83611	4.408	0.471 #
24)	L5 AR-1248-4	5.559	5.183	5265360	231546	10.363	1.111 #
25)	L5 AR-1248-5	5.643f	5.587	13646084	3187119	39.034	14.840 #
26)	L6 AR-1254-1	5.559	5.533	5265360	423661	9.592	1.289 #
27)	L6 AR-1254-2	5.706	5.679	4251496	674169	8.749	2.329 #
28)	L6 AR-1254-3	6.115	6.083	960417	390450	1.290	0.913 #
29)	L6 AR-1254-4	6.340	6.315	889201	272816	1.618	0.999 #
30)	L6 AR-1254-5	6.777	6.715	2127864	521981	3.006	1.563 #
31)	L7 AR-1260-1	6.243	6.215	1230642	367915	2.801	1.480 #
32)	L7 AR-1260-2	6.405	6.401	5161259	119185	7.503	0.362 #
33)	L7 AR-1260-3	6.817	6.561	606847	61403	1.032	0.239 #
34)	L7 AR-1260-4	7.040	7.009	2841469	546243	6.300	2.855 #
35)	L7 AR-1260-5	7.290	7.255	90666	1026632	0.073	2.491 #
36)	L8 AR-1262-1	6.817	6.752	606847	473336	0.653	1.141 #
37)	L8 AR-1262-2	7.290	7.255	90666	1026632	0.061	1.991 #
38)	L8 AR-1262-3	7.588	7.544	3597296	245098	6.095	1.314 #
39)	L8 AR-1262-4	7.639	7.628	2550814	145406	2.566	0.505 #
40)	L8 AR-1262-5	8.145	8.090	167107	294707	0.385	2.962 #
41)	L9 AR-1268-1	7.588	7.544	3597296	245098	2.054	0.473 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081225\
 Data File : P0112854.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Aug 2025 10:05
 Operator : YP/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 12 11:05:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 24 04:54:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	7.664	7.628	2081497	145406	1.413	0.342 #
43) L9	AR-1268-3	7.832	7.764f	1183235	14757	0.945	0.046 #
44) L9	AR-1268-4	8.145	8.090	167107	294707	0.347	2.648 #
45) L9	AR-1268-5	8.428	8.380	745402	87143	0.222	0.119 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081225\
Data File : PO112854.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2025 10:05
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 12 11:05:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 24 04:54:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.664 min
Delta R.T.: -0.005 min
Response: 170038746
Conc: 20.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.658 min
Delta R.T.: -0.004 min
Response: 101647548
Conc: 20.46 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.685 min
Delta R.T.: -0.008 min
Response: 125986606
Conc: 17.23 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.632 min
Delta R.T.: -0.008 min
Response: 36010722
Conc: 20.38 ng/ml



#3 AR-1016-1


R.T.: 4.818 min
Delta R.T.: 0.062 min
Response: 207474
Conc: 0.77 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#3 AR-1016-1

R.T.: 4.737 min
Delta R.T.: 0.000 min
Response: 147142
Conc: 0.84 ng/ml




#4 AR-1016-2

R.T.: 4.818 min
Delta R.T.: 0.044 min
Response: 207474
Conc: 0.51 ng/ml



#4 AR-1016-2


R.T.: 4.748 min
Delta R.T.: -0.008 min
Response: 236047
Conc: 0.91 ng/ml



#5 AR-1016-3


R.T.: 4.818 min
Delta R.T.: -0.013 min
Response: 207474
Conc: 0.79 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#5 AR-1016-3

R.T.: 4.932 min
Delta R.T.: 0.001 min
Response: 126024
Conc: 0.93 ng/ml




#6 AR-1016-4

R.T.: 4.945 min
Delta R.T.: -0.005 min
Response: 58929
Conc: 0.28 ng/ml



#6 AR-1016-4


R.T.: 4.975 min
Delta R.T.: 0.003 min
Response: 128539
Conc: 1.16 ng/ml



#7 AR-1016-5


R.T.: 5.198 min
Delta R.T.: -0.010 min
Response: 1515050
Conc: 6.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#7 AR-1016-5

R.T.: 5.183 min
Delta R.T.: -0.002 min
Response: 231546
Conc: 1.62 ng/ml




#8 AR-1221-1

R.T.: 3.867 min
Delta R.T.: -0.014 min
Response: 15226
Conc: 0.16 ng/ml



#8 AR-1221-1


R.T.: 3.875 min
Delta R.T.: 0.003 min
Response: 156746
Conc: 2.33 ng/ml



#9 AR-1221-2


R.T.: 3.967 min
Delta R.T.: 0.000 min
Response: 2377395
Conc: 33.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#9 AR-1221-2

R.T.: 3.957 min
Delta R.T.: 0.000 min
Response: 1699978
Conc: 33.51 ng/ml




#10 AR-1221-3

R.T.: 4.062 min
Delta R.T.: 0.018 min
Response: 258883
Conc: 1.09 ng/ml



#10 AR-1221-3


R.T.: 4.034 min
Delta R.T.: 0.003 min
Response: 46668
Conc: 0.29 ng/ml



#11 AR-1232-1


R.T.: 4.062 min
Delta R.T.: 0.019 min
Response: 258883
Conc: 1.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#11 AR-1232-1

R.T.: 4.034 min
Delta R.T.: 0.003 min
Response: 46668
Conc: 0.37 ng/ml




#12 AR-1232-2

R.T.: 4.531 min
Delta R.T.: -0.004 min
Response: 707238
Conc: 6.57 ng/ml



#12 AR-1232-2


R.T.: 4.748 min
Delta R.T.: -0.008 min
Response: 236047
Conc: 1.80 ng/ml



#13 AR-1232-3


R.T.: 4.818 min
Delta R.T.: 0.043 min
Response: 207474
Conc: 1.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#13 AR-1232-3

R.T.: 4.932 min
Delta R.T.: 0.001 min
Response: 126024
Conc: 1.88 ng/ml




#14 AR-1232-4

R.T.: 4.945 min
Delta R.T.: -0.005 min
Response: 58929
Conc: 0.56 ng/ml



#14 AR-1232-4


R.T.: 5.019 min
Delta R.T.: 0.004 min
Response: 83611
Conc: 1.40 ng/ml



#15 AR-1232-5


R.T.: 4.995 min
Delta R.T.: 0.000 min
Response: 247652
Conc: 3.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#15 AR-1232-5

R.T.: 5.183 min
Delta R.T.: -0.002 min
Response: 231546
Conc: 3.55 ng/ml




#16 AR-1242-1

R.T.: 4.818 min
Delta R.T.: 0.063 min
Response: 207474
Conc: 0.84 ng/ml



#16 AR-1242-1


R.T.: 4.737 min
Delta R.T.: 0.000 min
Response: 147142
Conc: 0.93 ng/ml



#17 AR-1242-2


R.T.: 4.818 min
Delta R.T.: 0.044 min
Response: 207474
Conc: 0.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#17 AR-1242-2

R.T.: 4.748 min
Delta R.T.: -0.008 min
Response: 236047
Conc: 1.01 ng/ml




#18 AR-1242-3

R.T.: 4.818 min
Delta R.T.: -0.013 min
Response: 207474
Conc: 0.88 ng/ml



#18 AR-1242-3


R.T.: 4.932 min
Delta R.T.: 0.001 min
Response: 126024
Conc: 1.04 ng/ml



#19 AR-1242-4


R.T.: 4.945 min
Delta R.T.: -0.005 min
Response: 58929
Conc: 0.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#19 AR-1242-4

R.T.: 5.019 min
Delta R.T.: 0.004 min
Response: 83611
Conc: 0.68 ng/ml




#20 AR-1242-5

R.T.: 5.643 min
Delta R.T.: 0.041 min
Response: 13646084
Conc: 63.79 ng/ml



#20 AR-1242-5


R.T.: 5.533 min
Delta R.T.: -0.001 min
Response: 423661
Conc: 2.63 ng/ml



#21 AR-1248-1


R.T.: 4.818 min
Delta R.T.: 0.062 min
Response: 207474
Conc: 1.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#21 AR-1248-1

R.T.: 4.737 min
Delta R.T.: 0.000 min
Response: 147142
Conc: 1.21 ng/ml




#22 AR-1248-2

R.T.: 4.995 min
Delta R.T.: 0.000 min
Response: 247652
Conc: 0.97 ng/ml



#22 AR-1248-2


R.T.: 4.975 min
Delta R.T.: 0.002 min
Response: 128539
Conc: 0.76 ng/ml



#23 AR-1248-3


R.T.: 5.198 min
Delta R.T.: -0.010 min
Response: 1515050
Conc: 4.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#23 AR-1248-3

R.T.: 5.019 min
Delta R.T.: 0.004 min
Response: 83611
Conc: 0.47 ng/ml




#24 AR-1248-4

R.T.: 5.559 min
Delta R.T.: -0.002 min
Response: 5265360
Conc: 10.36 ng/ml



#24 AR-1248-4


R.T.: 5.183 min
Delta R.T.: -0.003 min
Response: 231546
Conc: 1.11 ng/ml



#25 AR-1248-5


R.T.: 5.643 min
Delta R.T.: 0.041 min
Response: 13646084
Conc: 39.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#25 AR-1248-5

R.T.: 5.587 min
Delta R.T.: 0.011 min
Response: 3187119
Conc: 14.84 ng/ml




#26 AR-1254-1

R.T.: 5.559 min
Delta R.T.: -0.002 min
Response: 5265360
Conc: 9.59 ng/ml



#26 AR-1254-1


R.T.: 5.533 min
Delta R.T.: -0.002 min
Response: 423661
Conc: 1.29 ng/ml



#27 AR-1254-2


R.T.: 5.706 min
Delta R.T.: -0.005 min
Response: 4251496
Conc: 8.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#27 AR-1254-2

R.T.: 5.679 min
Delta R.T.: -0.003 min
Response: 674169
Conc: 2.33 ng/ml




#28 AR-1254-3

R.T.: 6.115 min
Delta R.T.: 0.002 min
Response: 960417
Conc: 1.29 ng/ml



#28 AR-1254-3


R.T.: 6.083 min
Delta R.T.: 0.000 min
Response: 390450
Conc: 0.91 ng/ml



#29 AR-1254-4


R.T.: 6.340 min
Delta R.T.: -0.003 min
Response: 889201
Conc: 1.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#29 AR-1254-4

R.T.: 6.315 min
Delta R.T.: 0.003 min
Response: 272816
Conc: 1.00 ng/ml




#30 AR-1254-5

R.T.: 6.777 min
Delta R.T.: 0.015 min
Response: 2127864
Conc: 3.01 ng/ml



#30 AR-1254-5


R.T.: 6.715 min
Delta R.T.: -0.012 min
Response: 521981
Conc: 1.56 ng/ml



#31 AR-1260-1


R.T.: 6.243 min
Delta R.T.: -0.002 min
Response: 1230642
Conc: 2.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#31 AR-1260-1

R.T.: 6.215 min
Delta R.T.: 0.002 min
Response: 367915
Conc: 1.48 ng/ml




#32 AR-1260-2

R.T.: 6.405 min
Delta R.T.: -0.029 min
Response: 5161259
Conc: 7.50 ng/ml



#32 AR-1260-2


R.T.: 6.401 min
Delta R.T.: -0.001 min
Response: 119185
Conc: 0.36 ng/ml



#33 AR-1260-3


R.T.: 6.817 min
Delta R.T.: 0.017 min
Response: 606847
Conc: 1.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#33 AR-1260-3

R.T.: 6.561 min
Delta R.T.: 0.007 min
Response: 61403
Conc: 0.24 ng/ml




#34 AR-1260-4

R.T.: 7.040 min
Delta R.T.: -0.019 min
Response: 2841469
Conc: 6.30 ng/ml



#34 AR-1260-4


R.T.: 7.009 min
Delta R.T.: -0.014 min
Response: 546243
Conc: 2.86 ng/ml



#35 AR-1260-5


R.T.: 7.290 min
Delta R.T.: -0.013 min
Response: 90666
Conc: 0.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#35 AR-1260-5

R.T.: 7.255 min
Delta R.T.: -0.009 min
Response: 1026632
Conc: 2.49 ng/ml




#36 AR-1262-1

R.T.: 6.817 min
Delta R.T.: 0.016 min
Response: 606847
Conc: 0.65 ng/ml



#36 AR-1262-1


R.T.: 6.752 min
Delta R.T.: -0.014 min
Response: 473336
Conc: 1.14 ng/ml



#37 AR-1262-2


R.T.: 7.290 min
Delta R.T.: -0.013 min
Response: 90666
Conc: 0.06 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#37 AR-1262-2

R.T.: 7.255 min
Delta R.T.: -0.010 min
Response: 1026632
Conc: 1.99 ng/ml




#38 AR-1262-3

R.T.: 7.588 min
Delta R.T.: 0.002 min
Response: 3597296
Conc: 6.09 ng/ml



#38 AR-1262-3


R.T.: 7.544 min
Delta R.T.: -0.002 min
Response: 245098
Conc: 1.31 ng/ml



#39 AR-1262-4


R.T.: 7.639 min
Delta R.T.: -0.011 min
Response: 2550814
Conc: 2.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#39 AR-1262-4

R.T.: 7.628 min
Delta R.T.: 0.018 min
Response: 145406
Conc: 0.50 ng/ml




#40 AR-1262-5

R.T.: 8.145 min
Delta R.T.: 0.000 min
Response: 167107
Conc: 0.38 ng/ml



#40 AR-1262-5


R.T.: 8.090 min
Delta R.T.: -0.012 min
Response: 294707
Conc: 2.96 ng/ml



#41 AR-1268-1


R.T.: 7.588 min
Delta R.T.: 0.003 min
Response: 3597296
Conc: 2.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#41 AR-1268-1

R.T.: 7.544 min
Delta R.T.: -0.002 min
Response: 245098
Conc: 0.47 ng/ml




#42 AR-1268-2

R.T.: 7.664 min
Delta R.T.: 0.012 min
Response: 2081497
Conc: 1.41 ng/ml



#42 AR-1268-2


R.T.: 7.628 min
Delta R.T.: 0.017 min
Response: 145406
Conc: 0.34 ng/ml



#43 AR-1268-3


R.T.: 7.832 min
Delta R.T.: -0.024 min
Response: 1183235
Conc: 0.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#43 AR-1268-3

R.T.: 7.764 min
Delta R.T.: -0.052 min
Response: 14757
Conc: 0.05 ng/ml




#44 AR-1268-4

R.T.: 8.145 min
Delta R.T.: 0.000 min
Response: 167107
Conc: 0.35 ng/ml



#44 AR-1268-4


R.T.: 8.090 min
Delta R.T.: -0.012 min
Response: 294707
Conc: 2.65 ng/ml



#45 AR-1268-5

R.T.: 8.428 min
Delta R.T.: -0.011 min
Response: 745402
Conc: 0.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK



#45 AR-1268-5

R.T.: 8.380 min
Delta R.T.: -0.011 min
Response: 87143
Conc: 0.12 ng/ml