

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081319\
 Data File : P0059243.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Aug 2019 14:56
 Operator : SM/SJ
 Sample : K4174-01DL 2X
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 VW-PAINT-CHIPSDL

Manual Integrations
 APPROVED

Ankita
 8/14/2019 2:47:53 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 13 15:15:35 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072819.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jul 27 06:49:05 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.150	3.510	405989	444683	9.620m	11.131m
2)	SA Decachlor...	9.649	8.387	663648	475807	12.254	12.054
Target Compounds							
26)	L6 AR-1254-1	6.143	5.348	779847	1426101	330.263m	550.504m#
27)	L6 AR-1254-2	6.361	5.494	1957833	784915	508.007m	342.843 #
28)	L6 AR-1254-3	6.727	5.890	1717767	1726153	424.178m	469.145m
29)	L6 AR-1254-4	7.010	6.117	1780705	1183931	494.552m	509.171m
30)	L6 AR-1254-5	7.425	6.526	2469340	1896492	681.242m	549.779m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081319\
Data File : PO059243.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Aug 2019 14:56
Operator : SM/SJ
Sample : K4174-01DL 2X
Misc :
ALS Vial : 11 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
VW-PAINT-CHIPSDL


Manual Integrations
APPROVED

Ankita
8/14/2019 2:47:53 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 13 15:15:35 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072819.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jul 27 06:49:05 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.150 min
Delta R.T.: -0.029 min
Response: 405989
Conc: 9.62 ng/ml m

Instrument :
ECD_O
ClientSampleId :
VW-PAINT-CHIPSDL


Manual Integrations
APPROVED

Ankita
8/14/2019 2:47:53 PM




#1 Tetrachloro-m-xylene

R.T.: 3.510 min
Delta R.T.: -0.014 min
Response: 444683
Conc: 11.13 ng/ml m




#2 Decachlorobiphenyl

R.T.: 9.649 min
Delta R.T.: -0.023 min
Response: 663648
Conc: 12.25 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.387 min
Delta R.T.: -0.027 min
Response: 475807
Conc: 12.05 ng/ml




#26 AR-1254-1

R.T.: 6.143 min
Delta R.T.: -0.030 min
Response: 779847
Conc: 330.26 ng/ml m

Instrument :
ECD_O
ClientSampleId :
VW-PAINT-CHIPSDL


Manual Integrations
APPROVED

Ankita
8/14/2019 2:47:53 PM




#26 AR-1254-1

R.T.: 5.348 min
Delta R.T.: -0.021 min
Response: 1426101
Conc: 550.50 ng/ml m




#27 AR-1254-2

R.T.: 6.361 min
Delta R.T.: -0.027 min
Response: 1957833
Conc: 508.01 ng/ml m



#27 AR-1254-2

R.T.: 5.494 min
Delta R.T.: -0.020 min
Response: 784915
Conc: 342.84 ng/ml




#28 AR-1254-3

R.T.: 6.727 min
Delta R.T.: -0.027 min
Response: 1717767
Conc: 424.18 ng/ml m

Instrument :
ECD_O
ClientSampleId :
VW-PAINT-CHIPSDL


Manual Integrations
APPROVED

Ankita
8/14/2019 2:47:53 PM




#28 AR-1254-3

R.T.: 5.890 min
Delta R.T.: -0.021 min
Response: 1726153
Conc: 469.14 ng/ml m




#29 AR-1254-4

R.T.: 7.010 min
Delta R.T.: -0.026 min
Response: 1780705
Conc: 494.55 ng/ml m



#29 AR-1254-4

R.T.: 6.117 min
Delta R.T.: -0.021 min
Response: 1183931
Conc: 509.17 ng/ml m




#30 AR-1254-5

R.T.: 7.425 min
Delta R.T.: -0.025 min
Response: 2469340
Conc: 681.24 ng/ml m

Instrument :
ECD_O
ClientSampleId :
VW-PAINT-CHIPSDL

Manual Integrations
APPROVED

Ankita
8/14/2019 2:47:53 PM



#30 AR-1254-5

R.T.: 6.526 min
Delta R.T.: -0.022 min
Response: 1896492
Conc: 549.78 ng/ml m