

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081423\
 Data File : P0097018.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Aug 2023 09:58
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 14 15:40:37 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 08 04:34:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.410	3.617	164.1E6	65000367	49.586	49.102
2) SA Decachlor...	10.347	8.633	119.2E6	55081949	54.465	47.610
Target Compounds						
21) L5 AR-1248-1	5.596	4.701	30816596	13921153	525.934	496.076
22) L5 AR-1248-2	5.875	4.939	46355810	19718155	493.055	465.260
23) L5 AR-1248-3	6.084	4.980	49025829	20710789	501.241	472.148
24) L5 AR-1248-4	6.496	5.152	48447998	24040323	533.313	470.918
25) L5 AR-1248-5	6.536	5.545	45940896	22370636	521.546	478.497



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081423\
Data File : PO097018.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Aug 2023 09:58
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 14 15:40:37 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO080723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 08 04:34:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.410 min
Delta R.T.: 0.003 min
Response: 164139720
Conc: 49.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.617 min
Delta R.T.: -0.004 min
Response: 65000367
Conc: 49.10 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.347 min
Delta R.T.: 0.008 min
Response: 119179521
Conc: 54.46 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.633 min
Delta R.T.: -0.011 min
Response: 55081949
Conc: 47.61 ng/ml



#21 AR-1248-1


R.T.: 5.596 min
Delta R.T.: 0.005 min
Response: 30816596
Conc: 525.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.701 min
Delta R.T.: -0.007 min
Response: 13921153
Conc: 496.08 ng/ml




#22 AR-1248-2

R.T.: 5.875 min
Delta R.T.: 0.006 min
Response: 46355810
Conc: 493.05 ng/ml



#22 AR-1248-2


R.T.: 4.939 min
Delta R.T.: -0.007 min
Response: 19718155
Conc: 465.26 ng/ml



#23 AR-1248-3


R.T.: 6.084 min
Delta R.T.: 0.007 min
Response: 49025829
Conc: 501.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.980 min
Delta R.T.: -0.008 min
Response: 20710789
Conc: 472.15 ng/ml




#24 AR-1248-4

R.T.: 6.496 min
Delta R.T.: 0.008 min
Response: 48447998
Conc: 533.31 ng/ml



#24 AR-1248-4


R.T.: 5.152 min
Delta R.T.: -0.007 min
Response: 24040323
Conc: 470.92 ng/ml



#25 AR-1248-5

R.T.: 6.536 min
Delta R.T.: 0.009 min
Response: 45940896
Conc: 521.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.545 min
Delta R.T.: -0.008 min
Response: 22370636
Conc: 478.50 ng/ml