

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081423\
 Data File : P0097063.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2023 00:57
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 15 06:43:37 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 08 04:34:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.410	3.616	149.1E6	63881632	45.048	48.257
2) SA Decachlor...	10.350	8.632	111.6E6	53849344	51.020	46.545
Target Compounds						
21) L5 AR-1248-1	5.595	4.701	28532524	12926832	486.952	460.644
22) L5 AR-1248-2	5.875	4.938	42250554	19341339	449.390	456.368
23) L5 AR-1248-3	6.083	4.980	46183820	20480762	472.184	466.904
24) L5 AR-1248-4	6.496	5.152	47027240	21639466	517.673	423.889
25) L5 AR-1248-5	6.536	5.545	45703245	21475228	518.848	459.345



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081423\
Data File : PO097063.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2023 00:57
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 15 06:43:37 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO080723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 08 04:34:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.410 min
Delta R.T.: 0.003 min
Response: 149117136
Conc: 45.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.616 min
Delta R.T.: -0.005 min
Response: 63881632
Conc: 48.26 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.350 min
Delta R.T.: 0.011 min
Response: 111641553
Conc: 51.02 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.632 min
Delta R.T.: -0.012 min
Response: 53849344
Conc: 46.55 ng/ml



#21 AR-1248-1


R.T.: 5.595 min
Delta R.T.: 0.004 min
Response: 28532524
Conc: 486.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.701 min
Delta R.T.: -0.007 min
Response: 12926832
Conc: 460.64 ng/ml




#22 AR-1248-2

R.T.: 5.875 min
Delta R.T.: 0.005 min
Response: 42250554
Conc: 449.39 ng/ml



#22 AR-1248-2


R.T.: 4.938 min
Delta R.T.: -0.008 min
Response: 19341339
Conc: 456.37 ng/ml



#23 AR-1248-3


R.T.: 6.083 min
Delta R.T.: 0.006 min
Response: 46183820
Conc: 472.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.980 min
Delta R.T.: -0.007 min
Response: 20480762
Conc: 466.90 ng/ml




#24 AR-1248-4

R.T.: 6.496 min
Delta R.T.: 0.008 min
Response: 47027240
Conc: 517.67 ng/ml



#24 AR-1248-4


R.T.: 5.152 min
Delta R.T.: -0.008 min
Response: 21639466
Conc: 423.89 ng/ml



#25 AR-1248-5

R.T.: 6.536 min
Delta R.T.: 0.008 min
Response: 45703245
Conc: 518.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.545 min
Delta R.T.: -0.008 min
Response: 21475228
Conc: 459.35 ng/ml