

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081423\
 Data File : P0097064.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2023 01:14
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 15 06:45:33 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 08 04:34:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.410	3.616	151.4E6	64765287	45.727	48.924
2) SA Decachlor...	10.351	8.632	113.8E6	54809952	52.025	47.375
Target Compounds						
26) L6 AR-1254-1	6.470	5.505	48675239	33645187	445.374	455.920
27) L6 AR-1254-2	6.695	5.653	72484088	30285501	449.576	459.705
28) L6 AR-1254-3	7.071	6.058	74322638	46343436	474.308	458.562
29) L6 AR-1254-4	7.364	6.287	51907803	28224014	514.866	487.674
30) L6 AR-1254-5	7.795	6.706	55834869	40210759	497.573	457.382



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081423\
Data File : PO097064.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2023 01:14
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 15 06:45:33 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO080723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 08 04:34:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.410 min
Delta R.T.: 0.003 min
Response: 151366439
Conc: 45.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.616 min
Delta R.T.: -0.005 min
Response: 64765287
Conc: 48.92 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.351 min
Delta R.T.: 0.011 min
Response: 113841033
Conc: 52.03 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.632 min
Delta R.T.: -0.012 min
Response: 54809952
Conc: 47.38 ng/ml



#26 AR-1254-1


R.T.: 6.470 min
Delta R.T.: 0.007 min
Response: 48675239
Conc: 445.37 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.505 min
Delta R.T.: -0.007 min
Response: 33645187
Conc: 455.92 ng/ml




#27 AR-1254-2

R.T.: 6.695 min
Delta R.T.: 0.008 min
Response: 72484088
Conc: 449.58 ng/ml



#27 AR-1254-2


R.T.: 5.653 min
Delta R.T.: -0.007 min
Response: 30285501
Conc: 459.70 ng/ml



#28 AR-1254-3


R.T.: 7.071 min
Delta R.T.: 0.010 min
Response: 74322638
Conc: 474.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.058 min
Delta R.T.: -0.008 min
Response: 46343436
Conc: 458.56 ng/ml




#29 AR-1254-4

R.T.: 7.364 min
Delta R.T.: 0.010 min
Response: 51907803
Conc: 514.87 ng/ml



#29 AR-1254-4


R.T.: 6.287 min
Delta R.T.: -0.007 min
Response: 28224014
Conc: 487.67 ng/ml



#30 AR-1254-5

R.T.: 7.795 min
Delta R.T.: 0.010 min
Response: 55834869
Conc: 497.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.706 min
Delta R.T.: -0.007 min
Response: 40210759
Conc: 457.38 ng/ml