

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081522\
 Data File : P0088961.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2022 11:40
 Operator : YP/AJ
 Sample : N4191-01 10X
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 2362

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/16/2022
 Supervised By :Ankita Jodhani 08/16/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 15 22:54:21 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 05 23:24:09 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.430	3.591	3232412	1646617	1.198m	1.521 #
2) SA Decachlor...	10.267	8.584	4608832	2082400	2.388	1.718 #

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081522\
Data File : PO088961.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2022 11:40
Operator : YP/AJ
Sample : N4191-01 10X
Misc :
ALS Vial : 9 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
2362


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/16/2022
Supervised By :Ankita Jodhani 08/16/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 15 22:54:21 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO080422.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 05 23:24:09 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.430 min
Delta R.T.: -0.009 min
Response: 3232412
Conc: 1.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
2362


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/16/2022
Supervised By :Ankita Jodhani 08/16/2022




#1 Tetrachloro-m-xylene

R.T.: 3.591 min
Delta R.T.: -0.007 min
Response: 1646617
Conc: 1.52 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.267 min
Delta R.T.: -0.010 min
Response: 4608832
Conc: 2.39 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.584 min
Delta R.T.: -0.008 min
Response: 2082400
Conc: 1.72 ng/ml