

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081522\
 Data File : P0089006.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Aug 2022 02:10
 Operator : YP/AJ
 Sample : N4202-01
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 SB10

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 16 06:25:14 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 05 23:24:09 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.432	3.591	59494702	21808720	22.057	20.150
2) SA Decachlor...	10.266	8.579	40703014	21975900	21.093	18.133

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081522\
Data File : PO089006.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Aug 2022 02:10
Operator : YP/AJ
Sample : N4202-01
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
SB10

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 16 06:25:14 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO080422.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 05 23:24:09 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.432 min
Delta R.T.: -0.007 min
Response: 59494702
Conc: 22.06 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB10




#1 Tetrachloro-m-xylene

R.T.: 3.591 min
Delta R.T.: -0.006 min
Response: 21808720
Conc: 20.15 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.266 min
Delta R.T.: -0.011 min
Response: 40703014
Conc: 21.09 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.579 min
Delta R.T.: -0.013 min
Response: 21975900
Conc: 18.13 ng/ml