

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081523\
 Data File : P0097077.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2023 12:30
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 16 01:45:42 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 08 04:34:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.410	3.616	156.7E6	65761033	47.325	49.676
2) SA Decachlor...	10.352	8.632	112.5E6	55938441	51.410	48.351
Target Compounds						
26) L6 AR-1254-1	6.471	5.505	50875337	34926896	465.505	473.289
27) L6 AR-1254-2	6.696	5.653	75264993	30937193	466.825	469.597
28) L6 AR-1254-3	7.073	6.057	74911020	47950881	478.063	474.467
29) L6 AR-1254-4	7.366	6.287	51598572	28961979	511.799	500.425
30) L6 AR-1254-5	7.797	6.706	56349439	41779050	502.159	475.220



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081523\
Data File : PO097077.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2023 12:30
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 16 01:45:42 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO080723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 08 04:34:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.410 min
Delta R.T.: 0.003 min
Response: 156655234
Conc: 47.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.616 min
Delta R.T.: -0.005 min
Response: 65761033
Conc: 49.68 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.352 min
Delta R.T.: 0.012 min
Response: 112494027
Conc: 51.41 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.632 min
Delta R.T.: -0.012 min
Response: 55938441
Conc: 48.35 ng/ml



#26 AR-1254-1


R.T.: 6.471 min
Delta R.T.: 0.008 min
Response: 50875337
Conc: 465.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.505 min
Delta R.T.: -0.007 min
Response: 34926896
Conc: 473.29 ng/ml




#27 AR-1254-2

R.T.: 6.696 min
Delta R.T.: 0.010 min
Response: 75264993
Conc: 466.82 ng/ml



#27 AR-1254-2


R.T.: 5.653 min
Delta R.T.: -0.006 min
Response: 30937193
Conc: 469.60 ng/ml



#28 AR-1254-3


R.T.: 7.073 min
Delta R.T.: 0.011 min
Response: 74911020
Conc: 478.06 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.057 min
Delta R.T.: -0.008 min
Response: 47950881
Conc: 474.47 ng/ml




#29 AR-1254-4

R.T.: 7.366 min
Delta R.T.: 0.011 min
Response: 51598572
Conc: 511.80 ng/ml



#29 AR-1254-4


R.T.: 6.287 min
Delta R.T.: -0.007 min
Response: 28961979
Conc: 500.42 ng/ml



#30 AR-1254-5

R.T.: 7.797 min
Delta R.T.: 0.012 min
Response: 56349439
Conc: 502.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.706 min
Delta R.T.: -0.007 min
Response: 41779050
Conc: 475.22 ng/ml