

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081523\
 Data File : P0097119.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Aug 2023 01:24
 Operator : YP/AJ
 Sample : PB154834BL
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB154834BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 16 05:46:02 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 08 04:34:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.409	3.615	65762877	25901560	19.867	19.566
2)	SA Decachlor...	10.348	8.630	47284616	23592531	21.609	20.392

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081523\
Data File : PO097119.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Aug 2023 01:24
Operator : YP/AJ
Sample : PB154834BL
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB154834BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 16 05:46:02 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO080723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 08 04:34:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.409 min
Delta R.T.: 0.002 min
Response: 65762877
Conc: 19.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB154834BL




#1 Tetrachloro-m-xylene

R.T.: 3.615 min
Delta R.T.: -0.006 min
Response: 25901560
Conc: 19.57 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.348 min
Delta R.T.: 0.008 min
Response: 47284616
Conc: 21.61 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.630 min
Delta R.T.: -0.014 min
Response: 23592531
Conc: 20.39 ng/ml