

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081524\
 Data File : P0105588.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Aug 2024 09:10
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/16/2024
 Supervised By :Ankita Jodhani 08/16/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 16 09:21:45 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081524.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Aug 15 11:44:23 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.454	3.726	520.5E6	144.9E6	55.671	56.795
2) SA Decachlor...	10.196	8.773	255.1E6	93917680	43.939	47.590
Target Compounds						
3) L1 AR-1016-1	5.617	4.819	148.0E6	49561112	475.435	568.302
4) L1 AR-1016-2	5.640	4.839	206.7E6	66413411	476.244	565.801
5) L1 AR-1016-3	5.702	5.016	125.2E6	35185845	470.881	569.169
6) L1 AR-1016-4	5.800	5.057	103.9E6	26551223	472.157	560.064
7) L1 AR-1016-5	6.094	5.272	105.6E6	35857277	505.765	564.649m
31) L7 AR-1260-1	7.218	6.310	157.3E6	64371530	425.766	543.413 #
32) L7 AR-1260-2	7.473	6.496	179.4E6	78693272	415.331	536.368 #
33) L7 AR-1260-3	7.832	6.651	109.5E6	75481547	404.214	491.383
34) L7 AR-1260-4	8.057	7.125	130.6E6	51679012	410.218	526.553 #
35) L7 AR-1260-5	8.377	7.363	263.2E6	123.0E6	422.555	508.227

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081524\
Data File : PO105588.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Aug 2024 09:10
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/16/2024
Supervised By :Ankita Jodhani 08/16/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 16 09:21:45 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO081524.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Aug 15 11:44:23 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

