

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081524\
 Data File : P0105552.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2024 14:36
 Operator : YP/AJ
 Sample : PB162750BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB162750BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 16 04:47:19 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081524.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Aug 15 11:44:23 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.452	3.725	214.9E6	58932870	22.983	23.106
2) SA Decachlor...	10.195	8.775	135.7E6	47609557	23.379	24.125

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081524\
Data File : PO105552.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2024 14:36
Operator : YP/AJ
Sample : PB162750BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB162750BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 16 04:47:19 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO081524.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Aug 15 11:44:23 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.452 min
Delta R.T.: 0.000 min
Response: 214876102
Conc: 22.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB162750BL




#1 Tetrachloro-m-xylene

R.T.: 3.725 min
Delta R.T.: -0.001 min
Response: 58932870
Conc: 23.11 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.195 min
Delta R.T.: 0.000 min
Response: 135708564
Conc: 23.38 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.775 min
Delta R.T.: -0.007 min
Response: 47609557
Conc: 24.12 ng/ml