

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081525\
 Data File : P0112984.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2025 15:28
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 01:47:53 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 24 04:54:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.665	3.658	460.2E6	266.7E6	56.590	53.673
2) SA Decachlor...	8.684	8.630	362.8E6	93070934	49.621	52.668
Target Compounds						
26) L6 AR-1254-1	5.554	5.528	286.2E6	159.6E6	521.383	485.680
27) L6 AR-1254-2	5.703	5.676	259.1E6	140.4E6	533.163	485.045
28) L6 AR-1254-3	6.106	6.076	387.0E6	209.8E6	519.963	490.881
29) L6 AR-1254-4	6.336	6.303	282.5E6	133.0E6	513.890	487.406
30) L6 AR-1254-5	6.754	6.719	367.8E6	168.8E6	519.674	505.420



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081525\
Data File : PO112984.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 15:28
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 01:47:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 24 04:54:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.665 min
Delta R.T.: -0.004 min
Response: 460163189
Conc: 56.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.658 min
Delta R.T.: -0.005 min
Response: 266709752
Conc: 53.67 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.684 min
Delta R.T.: -0.008 min
Response: 362821068
Conc: 49.62 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.630 min
Delta R.T.: -0.011 min
Response: 93070934
Conc: 52.67 ng/ml



#26 AR-1254-1


R.T.: 5.554 min
Delta R.T.: -0.007 min
Response: 286217626
Conc: 521.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.528 min
Delta R.T.: -0.007 min
Response: 159604611
Conc: 485.68 ng/ml




#27 AR-1254-2

R.T.: 5.703 min
Delta R.T.: -0.007 min
Response: 259073968
Conc: 533.16 ng/ml



#27 AR-1254-2


R.T.: 5.676 min
Delta R.T.: -0.007 min
Response: 140415970
Conc: 485.04 ng/ml



#28 AR-1254-3


R.T.: 6.106 min
Delta R.T.: -0.007 min
Response: 387001365
Conc: 519.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.076 min
Delta R.T.: -0.008 min
Response: 209832586
Conc: 490.88 ng/ml




#29 AR-1254-4

R.T.: 6.336 min
Delta R.T.: -0.007 min
Response: 282487461
Conc: 513.89 ng/ml



#29 AR-1254-4


R.T.: 6.303 min
Delta R.T.: -0.008 min
Response: 133043811
Conc: 487.41 ng/ml



#30 AR-1254-5

R.T.: 6.754 min
Delta R.T.: -0.008 min
Response: 367837728
Conc: 519.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.719 min
Delta R.T.: -0.008 min
Response: 168820055
Conc: 505.42 ng/ml