

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_O\Data\P0081619\  
 Data File : P0059368.D  
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
 Acq On : 16 Aug 2019 10:24  
 Operator : SM/SJ  
 Sample : K4356-07RE  
 Misc :  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 ECD\_O  
 ClientSampleId :  
 TP-4RE

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Aug 16 23:03:17 2019  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_O\methods\P0072819.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Wed Aug 14 09:07:26 2019  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound                    | RT#1  | RT#2  | Resp#1 | Resp#2 | ng/ml  | ng/ml   |
|-----------------------------|-------|-------|--------|--------|--------|---------|
| -----                       |       |       |        |        |        |         |
| System Monitoring Compounds |       |       |        |        |        |         |
| 1) SA Tetrachlo...          | 4.148 | 3.509 | 590547 | 549905 | 13.993 | 13.765  |
| 2) SA Decachlor...          | 9.641 | 8.383 | 274888 | 259458 | 5.076  | 6.573 # |

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_O\Data\P0081619\  
Data File : P0059368.D  
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
Acq On : 16 Aug 2019 10:24  
Operator : SM/SJ  
Sample : K4356-07RE  
Misc :  
ALS Vial : 8 Sample Multiplier: 1

Instrument :  
ECD\_O  
ClientSampleId :  
TP-4RE

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Aug 16 23:03:17 2019  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_O\methods\P0072819.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Wed Aug 14 09:07:26 2019  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.148 min  
Delta R.T.: 0.005 min  
Response: 590547  
Conc: 13.99 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
TP-4RE




#1 Tetrachloro-m-xylene

R.T.: 3.509 min  
Delta R.T.: 0.000 min  
Response: 549905  
Conc: 13.76 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.641 min  
Delta R.T.: 0.004 min  
Response: 274888  
Conc: 5.08 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.383 min  
Delta R.T.: 0.000 min  
Response: 259458  
Conc: 6.57 ng/ml