

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081621\
 Data File : P0080471.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Aug 2021 15:16
 Operator : DD\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 17 01:51:30 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 11 05:48:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.993	4.108	3874686	1263700	55.956	50.703
2) SA Decachlor...	11.096	9.507	2649909	1031358	51.193	45.471
Target Compounds						
26) L6 AR-1254-1	7.237	6.267	1282390	721699	494.690	491.932
27) L6 AR-1254-2	7.468	6.428	1987357	633260	506.453	490.211
28) L6 AR-1254-3	7.851	6.855	2078476	965757	509.248	478.967
29) L6 AR-1254-4	8.149	7.097	1525342	629383	510.912	483.618
30) L6 AR-1254-5	8.578	7.533	1586625	846995	512.807	482.774



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081621\
Data File : PO080471.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 Aug 2021 15:16
Operator : DD\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 17 01:51:30 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO081021.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 11 05:48:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.993 min
Delta R.T.: 0.000 min
Response: 3874686
Conc: 55.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.108 min
Delta R.T.: 0.000 min
Response: 1263700
Conc: 50.70 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.096 min
Delta R.T.: 0.000 min
Response: 2649909
Conc: 51.19 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.507 min
Delta R.T.: -0.005 min
Response: 1031358
Conc: 45.47 ng/ml



#26 AR-1254-1


R.T.: 7.237 min
Delta R.T.: -0.003 min
Response: 1282390
Conc: 494.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 6.267 min
Delta R.T.: -0.003 min
Response: 721699
Conc: 491.93 ng/ml




#27 AR-1254-2

R.T.: 7.468 min
Delta R.T.: -0.002 min
Response: 1987357
Conc: 506.45 ng/ml



#27 AR-1254-2


R.T.: 6.428 min
Delta R.T.: -0.002 min
Response: 633260
Conc: 490.21 ng/ml



#28 AR-1254-3


R.T.: 7.851 min
Delta R.T.: -0.002 min
Response: 2078476
Conc: 509.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.855 min
Delta R.T.: -0.003 min
Response: 965757
Conc: 478.97 ng/ml




#29 AR-1254-4

R.T.: 8.149 min
Delta R.T.: -0.001 min
Response: 1525342
Conc: 510.91 ng/ml



#29 AR-1254-4


R.T.: 7.097 min
Delta R.T.: -0.002 min
Response: 629383
Conc: 483.62 ng/ml



#30 AR-1254-5

R.T.: 8.578 min
Delta R.T.: -0.003 min
Response: 1586625
Conc: 512.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.533 min
Delta R.T.: -0.002 min
Response: 846995
Conc: 482.77 ng/ml