

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081622\
 Data File : P0089017.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Aug 2022 10:07
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 16 23:53:42 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 05 23:24:09 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.431	3.590	146.7E6	53176540	54.405	49.133
2) SA Decachlor...	10.271	8.580	111.3E6	60910696	57.683	50.260
Target Compounds						
26) L6 AR-1254-1	6.469	5.466	53375057	37003702	584.311	537.263
27) L6 AR-1254-2	6.689	5.614	79232398	31838648	579.382	537.632
28) L6 AR-1254-3	7.058	6.016	78907678	47915487	560.718	515.285
29) L6 AR-1254-4	7.344	6.245	57576245	29540980	569.808	521.459
30) L6 AR-1254-5	7.766	6.662	58680717	42586480	551.048	521.720



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081622\
Data File : PO089017.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Aug 2022 10:07
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 16 23:53:42 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO080422.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 05 23:24:09 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.431 min
Delta R.T.: -0.008 min
Response: 146749216
Conc: 54.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.590 min
Delta R.T.: -0.007 min
Response: 53176540
Conc: 49.13 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.271 min
Delta R.T.: -0.006 min
Response: 111309876
Conc: 57.68 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.580 min
Delta R.T.: -0.012 min
Response: 60910696
Conc: 50.26 ng/ml



#26 AR-1254-1


R.T.: 6.469 min
Delta R.T.: -0.005 min
Response: 53375057
Conc: 584.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.466 min
Delta R.T.: -0.009 min
Response: 37003702
Conc: 537.26 ng/ml




#27 AR-1254-2

R.T.: 6.689 min
Delta R.T.: -0.005 min
Response: 79232398
Conc: 579.38 ng/ml



#27 AR-1254-2


R.T.: 5.614 min
Delta R.T.: -0.009 min
Response: 31838648
Conc: 537.63 ng/ml



#28 AR-1254-3


R.T.: 7.058 min
Delta R.T.: -0.005 min
Response: 78907678
Conc: 560.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.016 min
Delta R.T.: -0.009 min
Response: 47915487
Conc: 515.28 ng/ml




#29 AR-1254-4

R.T.: 7.344 min
Delta R.T.: -0.004 min
Response: 57576245
Conc: 569.81 ng/ml



#29 AR-1254-4


R.T.: 6.245 min
Delta R.T.: -0.008 min
Response: 29540980
Conc: 521.46 ng/ml



#30 AR-1254-5

R.T.: 7.766 min
Delta R.T.: -0.004 min
Response: 58680717
Conc: 551.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.662 min
Delta R.T.: -0.008 min
Response: 42586480
Conc: 521.72 ng/ml