

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081622\
 Data File : P0089020.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Aug 2022 12:53
 Operator : YP/AJ
 Sample : N4211-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 WC-1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 16 23:56:47 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 05 23:24:09 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.431	3.588	61492066	20840964	22.797	19.256
2) SA Decachlor...	10.275	8.580	47873350	25062538	24.809	20.680

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081622\
Data File : PO089020.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Aug 2022 12:53
Operator : YP/AJ
Sample : N4211-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
WC-1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 16 23:56:47 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO080422.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 05 23:24:09 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.431 min
Delta R.T.: -0.008 min
Response: 61492066
Conc: 22.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
WC-1




#1 Tetrachloro-m-xylene

R.T.: 3.588 min
Delta R.T.: -0.009 min
Response: 20840964
Conc: 19.26 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.275 min
Delta R.T.: -0.002 min
Response: 47873350
Conc: 24.81 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.580 min
Delta R.T.: -0.012 min
Response: 25062538
Conc: 20.68 ng/ml