

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081624\
 Data File : P0105606.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Aug 2024 17:25
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 17 05:46:51 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081524.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Aug 15 11:44:23 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.454	3.726	473.8E6	132.5E6	50.677	51.961
2) SA Decachlor...	10.199	8.772	262.1E6	100.9E6	45.156	51.103
Target Compounds						
3) L1 AR-1016-1	5.617	4.819	160.5E6	45657940	515.749	523.545
4) L1 AR-1016-2	5.640	4.839	224.0E6	60695814	516.232	517.090
5) L1 AR-1016-3	5.703	5.017	135.5E6	31969282	509.711	517.138
6) L1 AR-1016-4	5.801	5.057	109.0E6	24417596	495.260	515.058
7) L1 AR-1016-5	6.095	5.273	102.5E6	31950240	491.074	503.125
31) L7 AR-1260-1	7.219	6.310	175.4E6	61990846	474.862	523.316
32) L7 AR-1260-2	7.475	6.496	198.8E6	76410210	460.314	520.807
33) L7 AR-1260-3	7.833	6.652	126.4E6	75298050	466.769	490.189
34) L7 AR-1260-4	8.059	7.125	148.4E6	51810790	466.042	527.896
35) L7 AR-1260-5	8.379	7.364	272.2E6	127.1E6	437.006	525.366



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081624\
Data File : PO105606.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Aug 2024 17:25
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 17 05:46:51 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO081524.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Aug 15 11:44:23 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.454 min
Delta R.T.: 0.001 min
Response: 473787160
Conc: 50.68 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.726 min
Delta R.T.: 0.000 min
Response: 132528840
Conc: 51.96 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.199 min
Delta R.T.: 0.003 min
Response: 262119784
Conc: 45.16 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.772 min
Delta R.T.: -0.010 min
Response: 100851682
Conc: 51.10 ng/ml



#3 AR-1016-1


R.T.: 5.617 min
Delta R.T.: 0.000 min
Response: 160530183
Conc: 515.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.819 min
Delta R.T.: -0.003 min
Response: 45657940
Conc: 523.55 ng/ml




#4 AR-1016-2

R.T.: 5.640 min
Delta R.T.: 0.000 min
Response: 224023628
Conc: 516.23 ng/ml



#4 AR-1016-2


R.T.: 4.839 min
Delta R.T.: -0.002 min
Response: 60695814
Conc: 517.09 ng/ml



#5 AR-1016-3


R.T.: 5.703 min
Delta R.T.: 0.000 min
Response: 135511911
Conc: 509.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.017 min
Delta R.T.: -0.003 min
Response: 31969282
Conc: 517.14 ng/ml




#6 AR-1016-4

R.T.: 5.801 min
Delta R.T.: 0.000 min
Response: 109020480
Conc: 495.26 ng/ml



#6 AR-1016-4


R.T.: 5.057 min
Delta R.T.: -0.003 min
Response: 24417596
Conc: 515.06 ng/ml



#7 AR-1016-5


R.T.: 6.095 min
Delta R.T.: 0.000 min
Response: 102536319
Conc: 491.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.273 min
Delta R.T.: -0.003 min
Response: 31950240
Conc: 503.12 ng/ml




#31 AR-1260-1

R.T.: 7.219 min
Delta R.T.: 0.000 min
Response: 175428744
Conc: 474.86 ng/ml



#31 AR-1260-1


R.T.: 6.310 min
Delta R.T.: -0.004 min
Response: 61990846
Conc: 523.32 ng/ml



#32 AR-1260-2


R.T.: 7.475 min
Delta R.T.: 0.001 min
Response: 198845427
Conc: 460.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.496 min
Delta R.T.: -0.004 min
Response: 76410210
Conc: 520.81 ng/ml




#33 AR-1260-3

R.T.: 7.833 min
Delta R.T.: 0.000 min
Response: 126419561
Conc: 466.77 ng/ml



#33 AR-1260-3


R.T.: 6.652 min
Delta R.T.: -0.005 min
Response: 75298050
Conc: 490.19 ng/ml



#34 AR-1260-4


R.T.: 8.059 min
Delta R.T.: 0.001 min
Response: 148381815
Conc: 466.04 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.125 min
Delta R.T.: -0.006 min
Response: 51810790
Conc: 527.90 ng/ml



#35 AR-1260-5

R.T.: 8.379 min
Delta R.T.: 0.002 min
Response: 272217235
Conc: 437.01 ng/ml



#35 AR-1260-5

R.T.: 7.364 min
Delta R.T.: -0.006 min
Response: 127101017
Conc: 525.37 ng/ml