

Data Path : Z:\bestpcbsrv\HPCHEM1\ECD 0\Data\P0081718\
 Data File : P0048243.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 17 Aug 2018 21:14
 Operator : SM/SJ
 Sample : J4511-07
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 RA-081518-NSW-45-029

Manual Integrations
 APPROVED

Sohil
 8/20/2018 3:53:23 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 03:29:02 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0072718.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jul 27 02:33:29 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.389	3.631	4847878	4914015	24.012	19.999m
2) SA Decachlor...	10.077	8.614	3068251	2645685	18.812	16.831
Target Compounds						
26) L6 AR-1254-1	6.627	5.513	65056915	50081217	5319.942	4069.983
27) L6 AR-1254-2	6.908	5.659	35544869	50842237	4766.129	4884.098
28) L6 AR-1254-3	6.993	6.061	88927488	110.4E6	6818.867	6362.627
29) L6 AR-1254-4	7.277	6.288	40885814	37197451	4194.971	3432.980m
30) L6 AR-1254-5	7.546	6.603	30617448	23102236	4144.333	3020.824 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD 0\Data\PO081718\
Data File : PO048243.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 17 Aug 2018 21:14
Operator : SM/SJ
Sample : J4511-07
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
RA-081518-NSW-45-029

Manual Integrations
APPROVED

Sohil
8/20/2018 3:53:23 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 03:29:02 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO072718.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jul 27 02:33:29 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

