

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081719\
 Data File : P0059431.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 17 Aug 2019 6:21
 Operator : SM/SJ
 Sample : K3616-01
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 JC-03-081519-A

Manual Integrations
 APPROVED

Ankita
 8/19/2019 3:33:54 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 17 23:24:22 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 17 03:09:44 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.142	3.509	940236	830071	26.854	23.060
2) SA Decachlor...	9.631	8.380	1114919	829234	20.957m	20.566

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081719\
Data File : PO059431.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 17 Aug 2019 6:21
Operator : SM/SJ
Sample : K3616-01
Misc :
ALS Vial : 52 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
JC-03-081519-A


Manual Integrations
APPROVED

Ankita
8/19/2019 3:33:54 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 17 23:24:22 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO081719.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 17 03:09:44 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.142 min
Delta R.T.: -0.007 min
Response: 940236
Conc: 26.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
JC-03-081519-A


Manual Integrations
APPROVED

Ankita
8/19/2019 3:33:54 PM




#1 Tetrachloro-m-xylene

R.T.: 3.509 min
Delta R.T.: -0.001 min
Response: 830071
Conc: 23.06 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.631 min
Delta R.T.: -0.010 min
Response: 1114919
Conc: 20.96 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.380 min
Delta R.T.: -0.003 min
Response: 829234
Conc: 20.57 ng/ml