

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_O\Data\P0081719\  
 Data File : P0059434.D  
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
 Acq On : 17 Aug 2019 7:09  
 Operator : SM/SJ  
 Sample : K4090-09  
 Misc :  
 ALS Vial : 55 Sample Multiplier: 1

Instrument :  
 ECD\_O  
 ClientSampleId :  
 OR-02-081519

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Aug 17 23:26:47 2019  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_O\methods\P0081719.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Sat Aug 17 03:09:44 2019  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound                    | RT#1  | RT#2  | Resp#1  | Resp#2 | ng/ml  | ng/ml  |
|-----------------------------|-------|-------|---------|--------|--------|--------|
| -----                       |       |       |         |        |        |        |
| System Monitoring Compounds |       |       |         |        |        |        |
| 1) SA Tetrachlo...          | 4.143 | 3.510 | 1049939 | 880017 | 29.988 | 24.448 |
| 2) SA Decachlor...          | 9.632 | 8.380 | 930573  | 652338 | 17.492 | 16.179 |

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_O\Data\PO081719\  
Data File : PO059434.D  
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
Acq On : 17 Aug 2019 7:09  
Operator : SM/SJ  
Sample : K4090-09  
Misc :  
ALS Vial : 55 Sample Multiplier: 1

Instrument :  
ECD\_O  
ClientSampleId :  
OR-02-081519

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Aug 17 23:26:47 2019  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_O\methods\PO081719.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Sat Aug 17 03:09:44 2019  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.143 min  
Delta R.T.: -0.006 min  
Response: 1049939  
Conc: 29.99 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
OR-02-081519




#1 Tetrachloro-m-xylene

R.T.: 3.510 min  
Delta R.T.: 0.000 min  
Response: 880017  
Conc: 24.45 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.632 min  
Delta R.T.: -0.009 min  
Response: 930573  
Conc: 17.49 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.380 min  
Delta R.T.: -0.004 min  
Response: 652338  
Conc: 16.18 ng/ml