

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081723\
 Data File : P0097224.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Aug 2023 23:48
 Operator : YP/AJ
 Sample : 04057-13
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 P001-SS061-0002-01

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 01:28:20 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 08 04:34:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.409	3.614	49247331	19876597	14.878	15.015
2)	SA Decachlor...	10.348	8.629	24234142	10506458	11.075	9.081
Target Compounds							
3)	L1 AR-1016-1	5.568	4.708	190490	220870	2.023	4.982 #
4)	L1 AR-1016-2	5.623	4.708	878812	220870	6.569	3.611 #
5)	L1 AR-1016-3	5.673	4.891	392554	139088	4.511	4.258
6)	L1 AR-1016-4	5.742	4.891f	443031	139088	6.596	4.866 #
7)	L1 AR-1016-5	6.086	5.136	97662	371554	1.399	9.664 #
8)	L2 AR-1221-1	4.619	0.000	1452481	0	35.639	N.D. #
9)	L2 AR-1221-2	4.719	3.908	824210	974532	27.130	75.424 #
10)	L2 AR-1221-3	4.775	4.035f	46949	773405	0.536	19.484 #
11)	L3 AR-1232-1	4.775	4.035f	46949	773405	0.644	23.910 #
12)	L3 AR-1232-2	5.339	4.708	1253951	220870	31.629	7.655 #
13)	L3 AR-1232-3	5.623	4.891	878812	139088	14.366	9.174 #
14)	L3 AR-1232-4	5.742	0.000	443031	0	14.519	N.D. #
15)	L3 AR-1232-5	5.837	5.136	-305	371554	N.D.	22.730 #
16)	L4 AR-1242-1	5.568	4.708	190490	220870	2.273	5.602 #
17)	L4 AR-1242-2	5.623	4.708	878812	220870	7.463	4.053 #
18)	L4 AR-1242-3	5.673	4.891	392554	139088	5.314	4.704
19)	L4 AR-1242-4	5.742	0.000	443031	0	7.442	N.D. #
20)	L4 AR-1242-5	6.561	5.499	-118425	387974	N.D.	10.648 #
21)	L5 AR-1248-1	5.568	4.708	190490	220870	3.251	7.871 #
22)	L5 AR-1248-2	5.837	4.891f	-305	139088	N.D.	3.282 #
23)	L5 AR-1248-3	6.086	0.000	97662	0	0.999	N.D. #
24)	L5 AR-1248-4	0.000	5.136	0	371554	N.D.	7.278 #
25)	L5 AR-1248-5	6.561	5.595f	-118425	1542321	N.D.	32.990 #
26)	L6 AR-1254-1	0.000	5.499	0	387974	N.D.	5.257 #
27)	L6 AR-1254-2	6.696	5.650	1032364	252549	6.403	3.833 #
28)	L6 AR-1254-3	7.072	6.072	2531063	6027333	16.153	59.640 #
29)	L6 AR-1254-4	7.365	6.283	1506429	201905	14.942	3.489 #
30)	L6 AR-1254-5	7.795	6.702	781726	946912	6.966	10.771 #
31)	L7 AR-1260-1	7.241	6.184	4329789	1093613	34.449	15.541 #
32)	L7 AR-1260-2	7.504	6.375	1116558	682975	8.287	8.183
33)	L7 AR-1260-3	7.876	6.526	2176064	220356	20.969	2.854 #
34)	L7 AR-1260-4	8.086	7.038	3541900	341719	31.680	5.333 #
35)	L7 AR-1260-5	8.428	7.246	4439728	314417	21.144	2.268 #
36)	L8 AR-1262-1	7.876	6.791	2176064	559982	14.651	14.224
37)	L8 AR-1262-2	8.428	7.038	4439728	341719	19.181	4.213 #
38)	L8 AR-1262-3	8.774	7.527	371664	329676	2.228	5.137 #
39)	L8 AR-1262-4	8.864	7.591	2491860	730490	34.179	6.430 #
40)	L8 AR-1262-5	9.588f	8.104	1008856	72696	12.114	1.384 #
41)	L9 AR-1268-1	8.774	7.527	371664	329676	1.209	1.779 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081723\
 Data File : P0097224.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Aug 2023 23:48
 Operator : YP/AJ
 Sample : 04057-13
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 P001-SS061-0002-01

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 01:28:20 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 08 04:34:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.864	7.591	2491860	730490	9.088	4.356 #
43) L9	AR-1268-3	9.098	7.801	1280109	762456	5.012	5.277
44) L9	AR-1268-4	9.588f	8.104	1008856	72696	10.881	1.220 #
45) L9	AR-1268-5	9.986	8.378	2384645	354244	2.952	0.852 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081723\
Data File : PO097224.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Aug 2023 23:48
Operator : YP/AJ
Sample : 04057-13
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
P001-SS061-0002-01

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 01:28:20 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO080723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 08 04:34:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.409 min
Delta R.T.: 0.002 min
Response: 49247331
Conc: 14.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-SS061-0002-01




#1 Tetrachloro-m-xylene

R.T.: 3.614 min
Delta R.T.: -0.007 min
Response: 19876597
Conc: 15.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.348 min
Delta R.T.: 0.008 min
Response: 24234142
Conc: 11.07 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.629 min
Delta R.T.: -0.016 min
Response: 10506458
Conc: 9.08 ng/ml



#3 AR-1016-1


R.T.: 5.568 min
Delta R.T.: -0.022 min
Response: 190490
Conc: 2.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-SS061-0002-01




#3 AR-1016-1

R.T.: 4.708 min
Delta R.T.: 0.000 min
Response: 220870
Conc: 4.98 ng/ml




#4 AR-1016-2

R.T.: 5.623 min
Delta R.T.: 0.009 min
Response: 878812
Conc: 6.57 ng/ml



#4 AR-1016-2


R.T.: 4.708 min
Delta R.T.: -0.018 min
Response: 220870
Conc: 3.61 ng/ml



#5 AR-1016-3


R.T.: 5.673 min
Delta R.T.: -0.004 min
Response: 392554
Conc: 4.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-SS061-0002-01




#5 AR-1016-3

R.T.: 4.891 min
Delta R.T.: -0.012 min
Response: 139088
Conc: 4.26 ng/ml




#6 AR-1016-4

R.T.: 5.742 min
Delta R.T.: -0.034 min
Response: 443031
Conc: 6.60 ng/ml



#6 AR-1016-4


R.T.: 4.891 min
Delta R.T.: -0.055 min
Response: 139088
Conc: 4.87 ng/ml



#7 AR-1016-5


R.T.: 6.086 min
Delta R.T.: 0.009 min
Response: 97662
Conc: 1.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-SS061-0002-01




#7 AR-1016-5

R.T.: 5.136 min
Delta R.T.: -0.004 min
Response: 371554
Conc: 9.66 ng/ml




#8 AR-1221-1

R.T.: 4.619 min
Delta R.T.: 0.001 min
Response: 1452481
Conc: 35.64 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T. : 3.835 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.719 min
Delta R.T.: 0.014 min
Response: 824210
Conc: 27.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-SS061-0002-01




#9 AR-1221-2

R.T.: 3.908 min
Delta R.T.: -0.013 min
Response: 974532
Conc: 75.42 ng/ml




#10 AR-1221-3

R.T.: 4.775 min
Delta R.T.: -0.007 min
Response: 46949
Conc: 0.54 ng/ml



#10 AR-1221-3


R.T.: 4.035 min
Delta R.T.: 0.039 min
Response: 773405
Conc: 19.48 ng/ml



#11 AR-1232-1


R.T.: 4.775 min
Delta R.T.: -0.007 min
Response: 46949
Conc: 0.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-SS061-0002-01




#11 AR-1232-1

R.T.: 4.035 min
Delta R.T.: 0.039 min
Response: 773405
Conc: 23.91 ng/ml




#12 AR-1232-2

R.T.: 5.339 min
Delta R.T.: 0.020 min
Response: 1253951
Conc: 31.63 ng/ml



#12 AR-1232-2


R.T.: 4.708 min
Delta R.T.: -0.018 min
Response: 220870
Conc: 7.66 ng/ml



#13 AR-1232-3


R.T.: 5.623 min
Delta R.T.: 0.010 min
Response: 878812
Conc: 14.37 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-SS061-0002-01




#13 AR-1232-3

R.T.: 4.891 min
Delta R.T.: -0.012 min
Response: 139088
Conc: 9.17 ng/ml




#14 AR-1232-4

R.T.: 5.742 min
Delta R.T.: -0.034 min
Response: 443031
Conc: 14.52 ng/ml



#14 AR-1232-4


R.T.: 0.000 min
Exp R.T.: 4.987 min
Response: 0
Conc: N.D.



#15 AR-1232-5


R.T.: 5.837 min
Delta R.T.: -0.032 min
Response: -305
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
P001-SS061-0002-01




#15 AR-1232-5

R.T.: 5.136 min
Delta R.T.: -0.023 min
Response: 371554
Conc: 22.73 ng/ml




#16 AR-1242-1

R.T.: 5.568 min
Delta R.T.: -0.023 min
Response: 190490
Conc: 2.27 ng/ml



#16 AR-1242-1


R.T.: 4.708 min
Delta R.T.: 0.000 min
Response: 220870
Conc: 5.60 ng/ml



#17 AR-1242-2


R.T.: 5.623 min
Delta R.T.: 0.010 min
Response: 878812
Conc: 7.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-SS061-0002-01




#17 AR-1242-2

R.T.: 4.708 min
Delta R.T.: -0.018 min
Response: 220870
Conc: 4.05 ng/ml




#18 AR-1242-3

R.T.: 5.673 min
Delta R.T.: -0.003 min
Response: 392554
Conc: 5.31 ng/ml



#18 AR-1242-3


R.T.: 4.891 min
Delta R.T.: -0.012 min
Response: 139088
Conc: 4.70 ng/ml



#19 AR-1242-4


R.T.: 5.742 min
Delta R.T.: -0.035 min
Response: 443031
Conc: 7.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-SS061-0002-01




#19 AR-1242-4

R.T.: 0.000 min
Exp R.T.: 4.987 min
Response: 0
Conc: N.D.




#20 AR-1242-5

R.T.: 6.561 min
Delta R.T.: 0.034 min
Response: -118425
Conc: N.D.



#20 AR-1242-5


R.T.: 5.499 min
Delta R.T.: -0.013 min
Response: 387974
Conc: 10.65 ng/ml



#21 AR-1248-1


R.T.: 5.568 min
Delta R.T.: -0.023 min
Response: 190490
Conc: 3.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-SS061-0002-01




#21 AR-1248-1

R.T.: 4.708 min
Delta R.T.: 0.000 min
Response: 220870
Conc: 7.87 ng/ml




#22 AR-1248-2

R.T.: 5.837 min
Delta R.T.: -0.033 min
Response: -305
Conc: N.D.



#22 AR-1248-2


R.T.: 4.891 min
Delta R.T.: -0.055 min
Response: 139088
Conc: 3.28 ng/ml



#23 AR-1248-3


R.T.: 6.086 min
Delta R.T.: 0.009 min
Response: 97662
Conc: 1.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-SS061-0002-01




#23 AR-1248-3

R.T.: 0.000 min
Exp R.T. : 4.987 min
Response: 0
Conc: N.D.




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T. : 6.488 min
Response: 0
Conc: N.D.



#24 AR-1248-4


R.T.: 5.136 min
Delta R.T.: -0.024 min
Response: 371554
Conc: 7.28 ng/ml



#25 AR-1248-5


R.T.: 6.561 min
Delta R.T.: 0.034 min
Response: -118425
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
P001-SS061-0002-01




#25 AR-1248-5

R.T.: 5.595 min
Delta R.T.: 0.041 min
Response: 1542321
Conc: 32.99 ng/ml




#26 AR-1254-1

R.T.: 0.000 min
Exp R.T.: 6.463 min
Response: 0
Conc: N.D.



#26 AR-1254-1


R.T.: 5.499 min
Delta R.T.: -0.012 min
Response: 387974
Conc: 5.26 ng/ml



#27 AR-1254-2


R.T.: 6.696 min
Delta R.T.: 0.010 min
Response: 1032364
Conc: 6.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-SS061-0002-01




#27 AR-1254-2

R.T.: 5.650 min
Delta R.T.: -0.009 min
Response: 252549
Conc: 3.83 ng/ml




#28 AR-1254-3

R.T.: 7.072 min
Delta R.T.: 0.011 min
Response: 2531063
Conc: 16.15 ng/ml



#28 AR-1254-3


R.T.: 6.072 min
Delta R.T.: 0.006 min
Response: 6027333
Conc: 59.64 ng/ml



#29 AR-1254-4


R.T.: 7.365 min
Delta R.T.: 0.010 min
Response: 1506429
Conc: 14.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-SS061-0002-01




#29 AR-1254-4

R.T.: 6.283 min
Delta R.T.: -0.012 min
Response: 201905
Conc: 3.49 ng/ml




#30 AR-1254-5

R.T.: 7.795 min
Delta R.T.: 0.011 min
Response: 781726
Conc: 6.97 ng/ml



#30 AR-1254-5


R.T.: 6.702 min
Delta R.T.: -0.011 min
Response: 946912
Conc: 10.77 ng/ml



#31 AR-1260-1


R.T.: 7.241 min
Delta R.T.: 0.011 min
Response: 4329789
Conc: 34.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-SS061-0002-01




#31 AR-1260-1

R.T.: 6.184 min
Delta R.T.: -0.012 min
Response: 1093613
Conc: 15.54 ng/ml




#32 AR-1260-2

R.T.: 7.504 min
Delta R.T.: 0.011 min
Response: 1116558
Conc: 8.29 ng/ml



#32 AR-1260-2


R.T.: 6.375 min
Delta R.T.: -0.011 min
Response: 682975
Conc: 8.18 ng/ml



#33 AR-1260-3


R.T.: 7.876 min
Delta R.T.: 0.012 min
Response: 2176064
Conc: 20.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-SS061-0002-01




#33 AR-1260-3

R.T.: 6.526 min
Delta R.T.: -0.014 min
Response: 220356
Conc: 2.85 ng/ml




#34 AR-1260-4

R.T.: 8.086 min
Delta R.T.: -0.010 min
Response: 3541900
Conc: 31.68 ng/ml



#34 AR-1260-4


R.T.: 7.038 min
Delta R.T.: 0.026 min
Response: 341719
Conc: 5.33 ng/ml



#35 AR-1260-5


R.T.: 8.428 min
Delta R.T.: -0.001 min
Response: 4439728
Conc: 21.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-SS061-0002-01




#35 AR-1260-5

R.T.: 7.246 min
Delta R.T.: -0.010 min
Response: 314417
Conc: 2.27 ng/ml




#36 AR-1262-1

R.T.: 7.876 min
Delta R.T.: 0.013 min
Response: 2176064
Conc: 14.65 ng/ml



#36 AR-1262-1


R.T.: 6.791 min
Delta R.T.: -0.013 min
Response: 559982
Conc: 14.22 ng/ml



#37 AR-1262-2


R.T.: 8.428 min
Delta R.T.: 0.000 min
Response: 4439728
Conc: 19.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-SS061-0002-01




#37 AR-1262-2

R.T.: 7.038 min
Delta R.T.: 0.027 min
Response: 341719
Conc: 4.21 ng/ml




#38 AR-1262-3

R.T.: 8.774 min
Delta R.T.: 0.017 min
Response: 371664
Conc: 2.23 ng/ml



#38 AR-1262-3


R.T.: 7.527 min
Delta R.T.: -0.012 min
Response: 329676
Conc: 5.14 ng/ml



#39 AR-1262-4


R.T.: 8.864 min
Delta R.T.: 0.014 min
Response: 2491860
Conc: 34.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-SS061-0002-01




#39 AR-1262-4

R.T.: 7.591 min
Delta R.T.: -0.012 min
Response: 730490
Conc: 6.43 ng/ml




#40 AR-1262-5

R.T.: 9.588 min
Delta R.T.: 0.046 min
Response: 1008856
Conc: 12.11 ng/ml



#40 AR-1262-5


R.T.: 8.104 min
Delta R.T.: 0.004 min
Response: 72696
Conc: 1.38 ng/ml



#41 AR-1268-1


R.T.: 8.774 min
Delta R.T.: 0.019 min
Response: 371664
Conc: 1.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-SS061-0002-01




#41 AR-1268-1

R.T.: 7.527 min
Delta R.T.: -0.012 min
Response: 329676
Conc: 1.78 ng/ml




#42 AR-1268-2

R.T.: 8.864 min
Delta R.T.: 0.009 min
Response: 2491860
Conc: 9.09 ng/ml



#42 AR-1268-2


R.T.: 7.591 min
Delta R.T.: -0.012 min
Response: 730490
Conc: 4.36 ng/ml



#43 AR-1268-3


R.T.: 9.098 min
Delta R.T.: 0.002 min
Response: 1280109
Conc: 5.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-SS061-0002-01




#43 AR-1268-3

R.T.: 7.801 min
Delta R.T.: -0.008 min
Response: 762456
Conc: 5.28 ng/ml




#44 AR-1268-4

R.T.: 9.588 min
Delta R.T.: 0.045 min
Response: 1008856
Conc: 10.88 ng/ml



#44 AR-1268-4


R.T.: 8.104 min
Delta R.T.: 0.004 min
Response: 72696
Conc: 1.22 ng/ml



#45 AR-1268-5

R.T.: 9.986 min
Delta R.T.: 0.003 min
Response: 2384645
Conc: 2.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-SS061-0002-01



#45 AR-1268-5

R.T.: 8.378 min
Delta R.T.: -0.011 min
Response: 354244
Conc: 0.85 ng/ml