

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081823\
 Data File : P0097283.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Aug 2023 00:11
 Operator : YP/AJ
 Sample : 04056-02
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 P001-PCBULK-001-02

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 21 04:40:24 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 08 04:34:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.409	3.614	65254529	26108184	19.713	19.722
2) SA Decachlor...	10.349	8.628	47689082	22762038	21.794	19.675
Target Compounds						
26) L6 AR-1254-1	6.470	5.501	225.1E6	173.3E6	2059.885	2348.501
27) L6 AR-1254-2	6.693	5.649	798.1E6	334.4E6	4950.301	5076.323
28) L6 AR-1254-3	7.071	6.053	1671.7E6	987.6E6	10668.281	9771.757
29) L6 AR-1254-4	7.365	6.283	1566.1E6	821.1E6	15534.218	14186.837
30) L6 AR-1254-5	7.795	6.702	2525.6E6	1680.7E6	22506.480	19117.096

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081823\
Data File : PO097283.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Aug 2023 00:11
Operator : YP/AJ
Sample : 04056-02
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
P001-PCBULK-001-02

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 21 04:40:24 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO080723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 08 04:34:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

