

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081920\
 Data File : P0070663.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 19 Aug 2020 9:03
 Operator : DD\AJ
 Sample : L3712-01
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 CHRT-27137

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 19 12:34:20 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080620.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Aug 06 13:04:06 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.378	3.558	1352008	832447	20.446	18.894
2) SA Decachlor...	10.001	8.741	2088073	1597079	24.386	28.772

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081920\
Data File : P0070663.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 19 Aug 2020 9:03
Operator : DD\AJ
Sample : L3712-01
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
CHRT-27137

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 19 12:34:20 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080620.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Aug 06 13:04:06 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.378 min
Delta R.T.: 0.004 min
Response: 1352008
Conc: 20.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
CHRT-27137




#1 Tetrachloro-m-xylene

R.T.: 3.558 min
Delta R.T.: 0.000 min
Response: 832447
Conc: 18.89 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.001 min
Delta R.T.: -0.004 min
Response: 2088073
Conc: 24.39 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.741 min
Delta R.T.: -0.012 min
Response: 1597079
Conc: 28.77 ng/ml