

Data Path : Z:\bestpcbsrv\HPCHEM1\ECD 0\Data\P0082018\
 Data File : P0048315.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Aug 2018 21:40
 Operator : SM/SJ
 Sample : J4525-09
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 21 01:10:33 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0072718.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jul 27 02:33:29 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.386	3.627	4492821	4407421	22.253	17.937
2) SA Decachlor...	10.068	8.607	2583079	2171207	15.837	13.812
Target Compounds						
3) L1 AR-1016-1	5.292	4.536	135329	858221	28.332	150.116 #
4) L1 AR-1016-2	5.650	4.906	22939	95908	3.897	18.266 #
5) L1 AR-1016-3	5.756	4.966	161175	272645	32.500	62.094 #
6) L1 AR-1016-4	6.059	5.023	199453	658369	42.881	126.965 #
7) L1 AR-1016-5	6.170	5.199	100704	48642	19.318	8.293 #
8) L2 AR-1221-1	4.594	3.864	186893	-31541	84.510	N.D. #
9) L2 AR-1221-2	4.711	3.944	280383	484834	171.467	267.265 #
10) L2 AR-1221-3	4.772	4.014	509463	648318	98.679	112.154
11) L3 AR-1232-1	5.101	4.324	424183	354882	197.246	150.902
12) L3 AR-1232-2	5.578	4.725	49230	182889	16.731	59.848 #
13) L3 AR-1232-3	5.578	4.749	49230	270243	11.459	58.522 #
14) L3 AR-1232-4	5.650	4.803	22939	275494	8.288	89.117 #
15) L3 AR-1232-5	5.756	4.966	161175	272645	72.174	152.339 #
16) L4 AR-1242-1	5.578	4.749	49230	270243	6.698	33.746 #
17) L4 AR-1242-2	5.650	4.906	22939	95908	4.956	23.281 #
18) L4 AR-1242-3	5.756	5.023	161175	658369	41.950	156.986 #
19) L4 AR-1242-4	6.059	5.199	199453	48642	51.883	10.301 #
20) L4 AR-1242-5	6.170	5.344	100704	156850	23.526	40.512 #
21) L5 AR-1248-1	6.059	5.199	199453	48642	31.897	6.520 #
22) L5 AR-1248-2	6.170	5.344	100704	156850	18.486	29.318 #
23) L5 AR-1248-3	6.406	5.530	387056	604635	55.000	60.765
24) L5 AR-1248-4	6.546	5.606	47111	314965	7.018	44.943 #
25) L5 AR-1248-5	6.633	6.094	146293	1333877	20.670	279.887 #
26) L6 AR-1254-1	6.633	5.530	146293	604635	11.963	49.137 #
27) L6 AR-1254-2	6.931	5.696	175009	28114	23.467	2.701 #
28) L6 AR-1254-3	6.996	6.094	669530	1333877	51.339	76.857 #
29) L6 AR-1254-4	7.289	6.308	588396	296048	60.371	27.323 #
30) L6 AR-1254-5	7.530	6.645	293448	94905	39.721	12.410 #
31) L7 AR-1260-1	7.175	6.197	186523	925524	19.891	83.757 #
32) L7 AR-1260-2	7.421	6.408	310562	143638	27.850	10.713 #
33) L7 AR-1260-3	7.709	6.718	94921	143329	7.378	9.859 #
34) L7 AR-1260-4	8.318	7.267	946055	99402	53.186	4.518 #
35) L7 AR-1260-5	8.653	7.637	589699	2977705	51.640	190.883 #
36) L8 AR-1262-1	7.175	6.408	186523	143638	22.515	12.669 #
37) L8 AR-1262-2	7.421	6.582	310562	265401	32.044	23.519 #
38) L8 AR-1262-3	7.766	6.749	104544	63552	8.519	4.211 #

Data Path : Z:\bestpcbsrv\HPCHEM1\ECD 0\Data\P0082018\
 Data File : P0048315.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Aug 2018 21:40
 Operator : SM/SJ
 Sample : J4525-09
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 21 01:10:33 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0072718.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jul 27 02:33:29 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
39) L8	AR-1262-4	7.969	7.045	778767	195867	66.133	14.874 #
40) L8	AR-1262-5	8.318	7.267	946055	99402	44.032	3.849 #
41) L9	AR-1268-1	8.602	7.537	804652	1599914	34.669	61.537 #
42) L9	AR-1268-2	8.715	7.637	929687	2977705	43.390	119.525 #
43) L9	AR-1268-3	8.942	7.829	563897	1924608	31.235	97.523 #
44) L9	AR-1268-4	9.389	8.071	825237	1462813	103.492	174.382 #
45) L9	AR-1268-5	9.750	8.362	469508	610176	8.617	11.176 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD 0\Data\PO082018\
Data File : PO048315.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Aug 2018 21:40
Operator : SM/SJ
Sample : J4525-09
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 21 01:10:33 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO072718.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jul 27 02:33:29 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.386 min
Delta R.T.: -0.012 min
Response: 4492821
Conc: 22.25 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.627 min
Delta R.T.: -0.019 min
Response: 4407421
Conc: 17.94 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.068 min
Delta R.T.: -0.021 min
Response: 2583079
Conc: 15.84 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.607 min
Delta R.T.: -0.038 min
Response: 2171207
Conc: 13.81 ng/ml



#3 AR-1016-1


R.T.: 5.292 min
Delta R.T.: -0.008 min
Response: 135329
Conc: 28.33 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.536 min
Delta R.T.: 0.025 min
Response: 858221
Conc: 150.12 ng/ml




#4 AR-1016-2

R.T.: 5.650 min
Delta R.T.: 0.000 min
Response: 22939
Conc: 3.90 ng/ml



#4 AR-1016-2


R.T.: 4.906 min
Delta R.T.: -0.020 min
Response: 95908
Conc: 18.27 ng/ml



#5 AR-1016-3


R.T.: 5.756 min
Delta R.T.: 0.007 min
Response: 161175
Conc: 32.50 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 4.966 min
Delta R.T.: -0.001 min
Response: 272645
Conc: 62.09 ng/ml




#6 AR-1016-4

R.T.: 6.059 min
Delta R.T.: 0.017 min
Response: 199453
Conc: 42.88 ng/ml



#6 AR-1016-4


R.T.: 5.023 min
Delta R.T.: 0.014 min
Response: 658369
Conc: 126.96 ng/ml



#7 AR-1016-5


R.T.: 6.170 min
Delta R.T.: -0.014 min
Response: 100704
Conc: 19.32 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.199 min
Delta R.T.: 0.018 min
Response: 48642
Conc: 8.29 ng/ml




#8 AR-1221-1

R.T.: 4.594 min
Delta R.T.: -0.014 min
Response: 186893
Conc: 84.51 ng/ml



#8 AR-1221-1


R.T.: 3.864 min
Delta R.T.: 0.003 min
Response: -31541
Conc: N.D.



#9 AR-1221-2


R.T.: 4.711 min
Delta R.T.: 0.017 min
Response: 280383
Conc: 171.47 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.944 min
Delta R.T.: -0.002 min
Response: 484834
Conc: 267.26 ng/ml




#10 AR-1221-3

R.T.: 4.772 min
Delta R.T.: 0.000 min
Response: 509463
Conc: 98.68 ng/ml



#10 AR-1221-3


R.T.: 4.014 min
Delta R.T.: -0.009 min
Response: 648318
Conc: 112.15 ng/ml



#11 AR-1232-1


R.T.: 5.101 min
Delta R.T.: -0.005 min
Response: 424183
Conc: 197.25 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 4.324 min
Delta R.T.: 0.004 min
Response: 354882
Conc: 150.90 ng/ml




#12 AR-1232-2

R.T.: 5.578 min
Delta R.T.: 0.012 min
Response: 49230
Conc: 16.73 ng/ml



#12 AR-1232-2


R.T.: 4.725 min
Delta R.T.: -0.005 min
Response: 182889
Conc: 59.85 ng/ml



#13 AR-1232-3


R.T.: 5.578 min
Delta R.T.: -0.010 min
Response: 49230
Conc: 11.46 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 4.749 min
Delta R.T.: 0.000 min
Response: 270243
Conc: 58.52 ng/ml




#14 AR-1232-4

R.T.: 5.650 min
Delta R.T.: 0.000 min
Response: 22939
Conc: 8.29 ng/ml



#14 AR-1232-4


R.T.: 4.803 min
Delta R.T.: -0.002 min
Response: 275494
Conc: 89.12 ng/ml



#15 AR-1232-5


R.T.: 5.756 min
Delta R.T.: 0.007 min
Response: 161175
Conc: 72.17 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 4.966 min
Delta R.T.: 0.000 min
Response: 272645
Conc: 152.34 ng/ml




#16 AR-1242-1

R.T.: 5.578 min
Delta R.T.: -0.010 min
Response: 49230
Conc: 6.70 ng/ml



#16 AR-1242-1


R.T.: 4.749 min
Delta R.T.: 0.000 min
Response: 270243
Conc: 33.75 ng/ml



#17 AR-1242-2


R.T.: 5.650 min
Delta R.T.: 0.000 min
Response: 22939
Conc: 4.96 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.906 min
Delta R.T.: -0.020 min
Response: 95908
Conc: 23.28 ng/ml




#18 AR-1242-3

R.T.: 5.756 min
Delta R.T.: 0.007 min
Response: 161175
Conc: 41.95 ng/ml



#18 AR-1242-3


R.T.: 5.023 min
Delta R.T.: 0.015 min
Response: 658369
Conc: 156.99 ng/ml



#19 AR-1242-4


R.T.: 6.059 min
Delta R.T.: 0.017 min
Response: 199453
Conc: 51.88 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 5.199 min
Delta R.T.: 0.019 min
Response: 48642
Conc: 10.30 ng/ml




#20 AR-1242-5

R.T.: 6.170 min
Delta R.T.: -0.013 min
Response: 100704
Conc: 23.53 ng/ml



#20 AR-1242-5


R.T.: 5.344 min
Delta R.T.: 0.015 min
Response: 156850
Conc: 40.51 ng/ml



#21 AR-1248-1


R.T.: 6.059 min
Delta R.T.: 0.017 min
Response: 199453
Conc: 31.90 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 5.199 min
Delta R.T.: 0.018 min
Response: 48642
Conc: 6.52 ng/ml




#22 AR-1248-2

R.T.: 6.170 min
Delta R.T.: -0.012 min
Response: 100704
Conc: 18.49 ng/ml



#22 AR-1248-2


R.T.: 5.344 min
Delta R.T.: 0.016 min
Response: 156850
Conc: 29.32 ng/ml



#23 AR-1248-3


R.T.: 6.406 min
Delta R.T.: -0.039 min
Response: 387056
Conc: 55.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 5.530 min
Delta R.T.: -0.001 min
Response: 604635
Conc: 60.76 ng/ml




#24 AR-1248-4

R.T.: 6.546 min
Delta R.T.: 0.064 min
Response: 47111
Conc: 7.02 ng/ml



#24 AR-1248-4


R.T.: 5.606 min
Delta R.T.: 0.034 min
Response: 314965
Conc: 44.94 ng/ml



#25 AR-1248-5


R.T.: 6.633 min
Delta R.T.: -0.007 min
Response: 146293
Conc: 20.67 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 6.094 min
Delta R.T.: 0.013 min
Response: 1333877
Conc: 279.89 ng/ml




#26 AR-1254-1

R.T.: 6.633 min
Delta R.T.: -0.003 min
Response: 146293
Conc: 11.96 ng/ml



#26 AR-1254-1


R.T.: 5.530 min
Delta R.T.: -0.002 min
Response: 604635
Conc: 49.14 ng/ml



#27 AR-1254-2


R.T.: 6.931 min
Delta R.T.: 0.015 min
Response: 175009
Conc: 23.47 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.696 min
Delta R.T.: 0.019 min
Response: 28114
Conc: 2.70 ng/ml




#28 AR-1254-3

R.T.: 6.996 min
Delta R.T.: -0.006 min
Response: 669530
Conc: 51.34 ng/ml



#28 AR-1254-3


R.T.: 6.094 min
Delta R.T.: 0.012 min
Response: 1333877
Conc: 76.86 ng/ml



#29 AR-1254-4


R.T.: 7.289 min
Delta R.T.: 0.003 min
Response: 588396
Conc: 60.37 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.308 min
Delta R.T.: 0.000 min
Response: 296048
Conc: 27.32 ng/ml




#30 AR-1254-5

R.T.: 7.530 min
Delta R.T.: -0.028 min
Response: 293448
Conc: 39.72 ng/ml



#30 AR-1254-5


R.T.: 6.645 min
Delta R.T.: 0.022 min
Response: 94905
Conc: 12.41 ng/ml



#31 AR-1260-1


R.T.: 7.175 min
Delta R.T.: 0.010 min
Response: 186523
Conc: 19.89 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.197 min
Delta R.T.: -0.014 min
Response: 925524
Conc: 83.76 ng/ml




#32 AR-1260-2

R.T.: 7.421 min
Delta R.T.: 0.000 min
Response: 310562
Conc: 27.85 ng/ml



#32 AR-1260-2


R.T.: 6.408 min
Delta R.T.: 0.011 min
Response: 143638
Conc: 10.71 ng/ml



#33 AR-1260-3


R.T.: 7.709 min
Delta R.T.: 0.006 min
Response: 94921
Conc: 7.38 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 6.718 min
Delta R.T.: -0.007 min
Response: 143329
Conc: 9.86 ng/ml




#34 AR-1260-4

R.T.: 8.318 min
Delta R.T.: 0.000 min
Response: 946055
Conc: 53.19 ng/ml



#34 AR-1260-4


R.T.: 7.267 min
Delta R.T.: 0.005 min
Response: 99402
Conc: 4.52 ng/ml



#35 AR-1260-5


R.T.: 8.653 min
Delta R.T.: 0.013 min
Response: 589699
Conc: 51.64 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.637 min
Delta R.T.: 0.028 min
Response: 2977705
Conc: 190.88 ng/ml




#36 AR-1262-1

R.T.: 7.175 min
Delta R.T.: 0.011 min
Response: 186523
Conc: 22.51 ng/ml



#36 AR-1262-1


R.T.: 6.408 min
Delta R.T.: 0.012 min
Response: 143638
Conc: 12.67 ng/ml



#37 AR-1262-2


R.T.: 7.421 min
Delta R.T.: 0.001 min
Response: 310562
Conc: 32.04 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 6.582 min
Delta R.T.: 0.029 min
Response: 265401
Conc: 23.52 ng/ml




#38 AR-1262-3

R.T.: 7.766 min
Delta R.T.: -0.012 min
Response: 104544
Conc: 8.52 ng/ml



#38 AR-1262-3


R.T.: 6.749 min
Delta R.T.: -0.014 min
Response: 63552
Conc: 4.21 ng/ml



#39 AR-1262-4


R.T.: 7.969 min
Delta R.T.: -0.035 min
Response: 778767
Conc: 66.13 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.045 min
Delta R.T.: 0.022 min
Response: 195867
Conc: 14.87 ng/ml




#40 AR-1262-5

R.T.: 8.318 min
Delta R.T.: 0.000 min
Response: 946055
Conc: 44.03 ng/ml



#40 AR-1262-5


R.T.: 7.267 min
Delta R.T.: 0.006 min
Response: 99402
Conc: 3.85 ng/ml



#41 AR-1268-1


R.T.: 8.602 min
Delta R.T.: -0.021 min
Response: 804652
Conc: 34.67 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.537 min
Delta R.T.: -0.008 min
Response: 1599914
Conc: 61.54 ng/ml




#42 AR-1268-2

R.T.: 8.715 min
Delta R.T.: 0.000 min
Response: 929687
Conc: 43.39 ng/ml



#42 AR-1268-2


R.T.: 7.637 min
Delta R.T.: 0.027 min
Response: 2977705
Conc: 119.53 ng/ml



#43 AR-1268-3


R.T.: 8.942 min
Delta R.T.: 0.004 min
Response: 563897
Conc: 31.23 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.829 min
Delta R.T.: 0.012 min
Response: 1924608
Conc: 97.52 ng/ml




#44 AR-1268-4

R.T.: 9.389 min
Delta R.T.: 0.031 min
Response: 825237
Conc: 103.49 ng/ml



#44 AR-1268-4


R.T.: 8.071 min
Delta R.T.: -0.030 min
Response: 1462813
Conc: 174.38 ng/ml



#45 AR-1268-5

R.T.: 9.750 min
Delta R.T.: -0.012 min
Response: 469508
Conc: 8.62 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.362 min
Delta R.T.: -0.027 min
Response: 610176
Conc: 11.18 ng/ml