

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0082020\
 Data File : P0070714.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Aug 2020 13:17
 Operator : DD\AJ
 Sample : P0082020ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 20 18:09:54 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0082020.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Aug 20 18:08:42 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	69.662	66.783 E3	4.1	97	0.00
2 SA Decachlorobiphenyl	93.621	84.895 E3	9.3	91	0.00
3 L1 AR-1016-1	2.751	2.611 E3	5.1	94	0.00
4 L1 AR-1016-2	4.050	3.781 E3	6.6	94	0.00
5 L1 AR-1016-3	2.371	2.289 E3	3.5	96	0.00
6 L1 AR-1016-4	1.981	1.944 E3	1.9	96	0.00
7 L1 AR-1016-5	1.942	1.886 E3	2.9	97	0.00
31 L7 AR-1260-1	4.122	3.910 E3	5.1	96	0.00
32 L7 AR-1260-2	6.216	5.897 E3	5.1	94	0.00
33 L7 AR-1260-3	5.359	5.044 E3	5.9	94	0.00
34 L7 AR-1260-4	5.091	4.740 E3	6.9	95	0.00
35 L7 AR-1260-5	11.914	11.083 E3	7.0	94	0.00

Signal #2

1 SA Tetrachloro-m-xylene	40.257	38.900 E3	3.4	99	0.00
2 SA Decachlorobiphenyl	56.312	54.701 E3	2.9	101	0.00
3 L1 AR-1016-1	1.725	1.669 E3	3.2	100	0.00
4 L1 AR-1016-2	2.411	2.314 E3	4.0	99	0.00
5 L1 AR-1016-3	1.241	1.240 E3	0.1	102	0.00
6 L1 AR-1016-4	0.973	1.036 E3	-6.5	103	0.00
7 L1 AR-1016-5	1.284	1.281 E3	0.2	102	0.00
31 L7 AR-1260-1	2.646	2.596 E3	1.9	101	0.00
32 L7 AR-1260-2	3.670	3.598 E3	2.0	101	0.00
33 L7 AR-1260-3	3.105	3.046 E3	1.9	100	0.00
34 L7 AR-1260-4	2.788	2.750 E3	1.4	101	0.00
35 L7 AR-1260-5	7.828	7.609 E3	2.8	100	0.00

Evaluate Continuing Calibration Report - Not Found

8 L2 AR-1221-1	672.379	0.000	100.0#	0#	-4.62#
9 L2 AR-1221-2	541.557	0.000	100.0#	0#	-4.71#
10 L2 AR-1221-3	1.782	0.000 E3	100.0#	0#	-4.80#
11 L3 AR-1232-1	1.463	0.000 E3	100.0#	0#	-4.80#
12 L3 AR-1232-2	825.946	0.000	100.0#	0#	-5.37#
13 L3 AR-1232-3	1.604	0.000 E3	100.0#	0#	-5.69#
14 L3 AR-1232-4	789.384	0.000	100.0#	0#	-5.86#
15 L3 AR-1232-5	513.950	0.000	100.0#	0#	-5.96#
16 L4 AR-1242-1	2.138	0.000 E3	100.0#	0#	-5.66#
17 L4 AR-1242-2	3.138	0.000 E3	100.0#	0#	-5.69#
18 L4 AR-1242-3	1.892	0.000 E3	100.0#	0#	-5.75#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0082020\
 Data File : P0070714.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Aug 2020 13:17
 Operator : DD\AJ
 Sample : P0082020ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 20 18:09:54 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0082020.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Aug 20 18:08:42 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
19 L4 AR-1242-4	1.616	0.000 E3	100.0#	0#	-5.86#
20 L4 AR-1242-5	2.092	0.000 E3	100.0#	0#	-6.63#
21 L5 AR-1248-1	1.910	0.000 E3	100.0#	0#	-5.69#
22 L5 AR-1248-2	2.100	0.000 E3	100.0#	0#	-5.95#
23 L5 AR-1248-3	2.502	0.000 E3	100.0#	0#	-6.16#
24 L5 AR-1248-4	3.509	0.000 E3	100.0#	0#	-6.59#
25 L5 AR-1248-5	3.263	0.000 E3	100.0#	0#	-6.63#
26 L6 AR-1254-1	3.343	0.000 E3	100.0#	0#	-6.56#
27 L6 AR-1254-2	5.263	0.000 E3	100.0#	0#	-6.79#
28 L6 AR-1254-3	5.437	0.000 E3	100.0#	0#	-7.16#
29 L6 AR-1254-4	4.922	0.000 E3	100.0#	0#	-7.46#
30 L6 AR-1254-5	5.288	0.000 E3	100.0#	0#	-7.88#
36 L8 AR-1262-1	7.651	0.000 E3	100.0#	0#	-7.95#
37 L8 AR-1262-2	12.982	0.000 E3	100.0#	0#	-8.49#
38 L8 AR-1262-3	5.615	0.000 E3	100.0#	0#	-8.76#
39 L8 AR-1262-4	3.246	0.000 E3	100.0#	0#	-8.84#
40 L8 AR-1262-5	4.083	0.000 E3	100.0#	0#	-9.39#
41 L9 AR-1268-1	15.634	0.000 E3	100.0#	0#	-8.76#
42 L9 AR-1268-2	13.352	0.000 E3	100.0#	0#	-8.84#
43 L9 AR-1268-3	11.510	0.000 E3	100.0#	0#	-9.02#
44 L9 AR-1268-4	5.072	0.000 E3	100.0#	0#	-9.39#
45 L9 AR-1268-5	33.809	0.000 E3	100.0#	0#	-9.72#

Signal #2

8 L2 AR-1221-1	476.599	0.000	100.0#	0#	-3.80#
9 L2 AR-1221-2	379.823	0.000	100.0#	0#	-3.89#
10 L2 AR-1221-3	1.150	0.000 E3	100.0#	0#	-3.98#
11 L3 AR-1232-1	880.515	0.000	100.0#	0#	-3.98#
12 L3 AR-1232-2	989.528	0.000	100.0#	0#	-4.79#
13 L3 AR-1232-3	535.362	0.000	100.0#	0#	-4.98#
14 L3 AR-1232-4	425.366	0.000	100.0#	0#	-5.08#
15 L3 AR-1232-5	493.801	0.000	100.0#	0#	-5.25#
16 L4 AR-1242-1	1.347	0.000 E3	100.0#	0#	-4.77#
17 L4 AR-1242-2	1.899	0.000 E3	100.0#	0#	-4.79#
18 L4 AR-1242-3	1.007	0.000 E3	100.0#	0#	-4.98#
19 L4 AR-1242-4	896.196	0.000	100.0#	0#	-5.07#
20 L4 AR-1242-5	1.353	0.000 E3	100.0#	0#	-5.63#
21 L5 AR-1248-1	1.076	0.000 E3	100.0#	0#	-4.79#
22 L5 AR-1248-2	1.352	0.000 E3	100.0#	0#	-5.03#
23 L5 AR-1248-3	1.270	0.000 E3	100.0#	0#	-5.07#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0082020\
 Data File : P0070714.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Aug 2020 13:17
 Operator : DD\AJ
 Sample : P0082020ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 20 18:09:54 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0082020.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Aug 20 18:08:42 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
24 L5 AR-1248-4	1.643	0.000 E3	100.0#	0#	-5.25#
25 L5 AR-1248-5	1.744	0.000 E3	100.0#	0#	-5.67#
26 L6 AR-1254-1	2.772	0.000 E3	100.0#	0#	-5.63#
27 L6 AR-1254-2	2.479	0.000 E3	100.0#	0#	-5.79#
28 L6 AR-1254-3	3.997	0.000 E3	100.0#	0#	-6.20#
29 L6 AR-1254-4	3.000	0.000 E3	100.0#	0#	-6.44#
30 L6 AR-1254-5	3.952	0.000 E3	100.0#	0#	-6.86#
36 L8 AR-1262-1	2.088	0.000 E3	100.0#	0#	-6.86#
37 L8 AR-1262-2	8.563	0.000 E3	100.0#	0#	-7.41#
38 L8 AR-1262-3	3.547	0.000 E3	100.0#	0#	-7.69#
39 L8 AR-1262-4	6.086	0.000 E3	100.0#	0#	-7.76#
40 L8 AR-1262-5	3.000	0.000 E3	100.0#	0#	-8.25#
41 L9 AR-1268-1	10.686	0.000 E3	100.0#	0#	-7.69#
42 L9 AR-1268-2	9.049	0.000 E3	100.0#	0#	-7.76#
43 L9 AR-1268-3	7.729	0.000 E3	100.0#	0#	-7.96#
44 L9 AR-1268-4	3.433	0.000 E3	100.0#	0#	-8.25#
45 L9 AR-1268-5	22.642	0.000 E3	100.0#	0#	-8.52#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0